Paper-based electrode assemble for impedimetric detection of miRNA

Eksin, E., Torul, H., Yarali, E., Tamer, U., Papakonstantinou, P., & Erdem, A. (2021). Paper-based electrode assemble for impedimetric detection of miRNA. *Talanta*, 225, 122043. [122043].
<https://doi.org/10.1016/j.talanta.2020.122043>

[Link to publication record in Ulster University Research Portal](#)

Published in:
Talanta

Publication Status:
Published (in print/issue): 01/04/2021

DOI:
[10.1016/j.talanta.2020.122043](https://doi.org/10.1016/j.talanta.2020.122043)

General rights

Copyright for the publications made accessible via Ulster University's Research Portal is retained by the author(s) and / or other copyright owners and it is a condition of accessing these publications that users recognise and abide by the legal requirements associated with these rights.

Take down policy

The Research Portal is Ulster University's institutional repository that provides access to Ulster's research outputs. Every effort has been made to ensure that content in the Research Portal does not infringe any person's rights, or applicable UK laws. If you discover content in the Research Portal that you believe breaches copyright or violates any law, please contact pure-support@ulster.ac.uk.

Paper-based electrode assemble for impedimetric detection of miRNA

Ece Eksin^{1,‡}, Hilal Torul^{2,‡}, Ece Yarali^{1,‡}, Ugur Tamer^{2,*}, Pagona Papakonstantinou³ and

Arzum Erdem^{1,*}

¹ Department of Analytical Chemistry, Faculty of Pharmacy, Ege University, Bornova, 35100 İzmir, Turkey

² Department of Analytical Chemistry, Faculty of Pharmacy, Gazi University, Etiler, 06330, Ankara, Turkey

³ School of Engineering, Engineering Research Institute, Ulster University, Newtownabbey BT37 0QB, United Kingdom

Supplementary Material

Sequences of synthetic oligonucleotides:

thiol link miRNA-155 DNA probe

5'- SH-ACC CCT ATC ACG ATT AGC ATT AA-3'

miRNA-155 RNA target

5'- UUA AUG CUA AUC GUG AUA GGG GU-3'

miRNA-155 Non-complementary (NC)

5'-UGG CAG UGU CUU AGC UGG UUG U-3'

miRNA-155 Mismatch (MM)

5'-UUA AUG CUA AUC GUC AUA GGG GU-3'

DNA probe and miRNA target stock solutions was prepared in Tris-EDTA buffer solution (pH 8.00) and kept frozen. The diluted solutions were prepared in PBS (pH 7.40).

Impedimetric measurements

The Randles circuit which consist of the solution resistance (R_s), the capacitance (Q), the charge transfer resistance (R_{ct}) and Warburg impedance (W), was used as equivalent circuit model used for fitting of the impedance data. The equivalent circuit model, which was given as inset in all Nyquist diagrams, used for fitting of the impedance data.

Fig. S1. The fabrication process of the paper electrode.

Fig. S2. The fabrication process of the paper electrode.

Fig. S3. Storage stability of the paper electrodes. The oxidation signal obtained by CV measurement in 50 mM $K_3[Fe(CN)_6]$ solution that prepared in 0.1 M KCl ($n=3$).

Table S1. The average values ($n=4$) of anodic peaks (I_a), cathodic peaks (I_c), relative anodic charge (Q_a), relative cathodic charge (Q_c) and calculated surface area were measured by CV technique for PE and AuNP-PWE.

Electrodes	I_a (μA)	I_c (μA)	Q_a (mC)	Q_c (mC)	A (cm^2)
PE	234.65 ± 12.51 (RSD%, 5.33%)	239.58 ± 15.36 (RSD%, 6.41%)	2.23	3.10	0.0198
AuNP-PE	447.53 ± 45.44 (RSD%, 10.15%)	430.10 ± 28.68 (RSD%, 6.67%)	2.56	4.66	0.0379

Fig. S4. Line graph representing the R_{ct} values recorded by the hybridization of 0.5 $\mu\text{g/mL}$ miRNA 155 DNA probe and miRNA 155 target with its various concentrations from 0 to 2 $\mu\text{g/mL}$ ($n=3$).

Table S2. The R_{ct} values ($n=3$) recorded by AuNP-PE after hybridization of probe with miRNA-155, NC, MM, or mixture samples and hybridization efficiency (HE %) calculated according to the R_{ct} values obtained after hybridization.

	Average R_{ct} values ($n=3$)	HE %
miRNA 155	31.63 ± 7.74 Ohm (RSD %, 24.46%)	100*
NC	19.80 ± 1.15 Ohm (RSD %, 5.82%)	43
MM	22.90 ± 3.61 Ohm (RSD %, 15.74)	58
miRNA 155 : NC mix	28.35 ± 5.20 Ohm (RSD %, 18.35%)	84
miRNA 155 : MM mix	23.40 ± 1.76 Ohm (RSD %, 7.51%)	60

*Accepted case

Fig. S5. Line graph representing the average R_{ct} values recorded by AuNP-PEs in case of hybridization of probe and miRNA 155 with its various concentrations from 0 to 6 $\mu\text{g/mL}$ in 1:400 diluted FBS medium.

Fig. S6. Selectivity of AuNP-PE in FBS medium. Histograms representing the average R_{ct} values obtained by miRNA 155 probe/AuNP-PE, hybridization between probe and miRNA 155, non-complementary (NC), single-base mismatched strand (MM), and mixture samples with miRNA 155 and NC or MM ($n=3$).

Table S3. The R_{ct} values (n=3) recorded by AuNP-PE after hybridization of probe with miRNA-155, NC, MM, or mixture samples in 1:400 diluted FBS medium and hybridization efficiency (HE %) calculated according to the R_{ct} values obtained after hybridization.

	Average R_{ct} values (n=3)	HE %
miRNA 155 target	83.75 ± 8.67 Ohm (RSD %, 10.36%)	100*
NC	43.43 ± 8.45 Ohm (RSD %, 19.46)	13
MM	44.63 ± 10.50 Ohm (RSD %, 23.52%)	16
miRNA 155 : NC mix	67.70 ± 11.10 Ohm (RSD %, 16.40%)	65
miRNA 155 : MM mix	56.20 ± 9.77 Ohm (RSD %, 17.39%)	41

*Accepted case

Table S4. Comparison of different biosensors for determination of miRNA. **Abbreviations:** IL: ionic liquid, CA: covalent agent, PGE: pencil graphite electrode, CNF: carbon nanofiber, PPy: polypyrrole, GO: graphene oxide, MWCNT: multiwalled carbon nanotube, GCE: glassy carbon electrode, HMDE: hanging mercury drop electrode, AuE: gold electrode, CB: carbon black, ITO: indium tin oxide, GONRs: graphene oxide nanoribbons, SPE: screen printed electrode, DPV: differential pulse voltammetry, SWV: square wave voltammetry, ASV: stripping voltammetry, CC: chronocoulometry.

miRNA	Electrode	Method	Indicator	Analysis Time	Real Sample	DL	Reference
miRNA-34a	IL-CA-PGE	DPV	-	35 min	+	0.56 $\mu\text{g/mL}$ (PBS) 0.40 $\mu\text{g/mL}$ (FBS)	[42]
	GO/CA/PGE	EIS	-	2 h	+	1.12 $\mu\text{g/mL}$ (PBS) 0.29 $\mu\text{g/mL}$ (FBS)	[43]
	CNFs/SPE	DPV, EIS	-	50 min	-	10.98 $\mu\text{g/mL}$	[44]
	PPy-PGE	EIS	-	40 min	+	0.20 $\mu\text{g/mL}$	[45]
	GO/CA/PGE	DPV	-	3.5 h	-	7.52 $\mu\text{g/mL}$	[46]
	IL/CA/PGE	EIS	-	2.5 h	+	0.772 $\mu\text{g/mL}$	[47]
	GO/PGE	CV, EIS	-	1.5 h	+	1.9 $\mu\text{g/mL}$	[48]
	MWCNT-GCE	DPV	-	1 h	+	1 pM	[49]
miRNA-522	Os(VI)bipy-HMDE	DPV	+	1.5 h	-	2 nM	[50]
miRNA-21 and miRNA-141	MoS ₂ /AuNPs/Ag NW paper electrode	SWV	+	17 h	+	0.1 fM	[31]
miRNA-21	MoS ₂ -Thi – AuNPs nanocomposite/ GCE	SWV	+	18 h	+	0.26 pM	[51]
	AuNPs/ITO	ASV	+	17 h	+	0.12 fM	[52]
	AuNPs/GCE	DPV	+	3.5 h	+	78 aM	[29]

	AuNPs@MoS ₂ /GCE	DPV EIS	+	17 h	+	0.78 fM 0.45 fM	[53]
	MWCNTs@GONRs/AuNPs	DPV	+	14 h	+	0.034 fM	[54]
	AuE	DPV	+	14 h	+	0.04 fM	[55]
	AuNPs@MoS ₂ (10k)/SPGE AuNPs@/SPGE	CC	+	22 h	+	100 aM 10 fM	[41]
let 7a	Gold disk electrode	DPV	+	25 h	-	1 nM	[56]
miRNA-145	AuE	EIS, SWV	+	18 h	-	0.37 fM	[57]
miRNA-125a	CB/PGE	EIS	-	2.5 min	+	0.01 nM	[58]
	MWCNTs/PGE					1 nM	
	GO/PGE					0.1 nM	
miRNA-155	AuE	SWV	+		+	0.13 fM	[59]
miRNA-122	GO/AuE	DPV	+	29 h	+	1.5 fM	[60]
miRNA-375	Au-SPE	SWV	+	4 h	+	11.7 aM	[61]
miRNA-155	AuNP-PE	EIS	-	15 min	+	33.8 nM (PBS) 93.4 nM (FBS)	This work