An audit of subcutaneous syringe drivers in a non-specialist hospital.

Dunne, K., Sullivan, K., Garvey, A., Kernohan, G., Diamond, A., Duffy, C., & Hutchinson, J. (2000). An audit of subcutaneous syringe drivers in a non-specialist hospital. *International Journal of Palliative Nursing*, 6(5), 214-219. <https://doi.org/10.12968/ijpn.2000.6.5.8922>

[Link to publication record in Ulster University Research Portal](#)

Published in:

International Journal of Palliative Nursing

Publication Status:

Published (in print/issue): 01/05/2000

DOI:

[10.12968/ijpn.2000.6.5.8922](https://doi.org/10.12968/ijpn.2000.6.5.8922)

Document Version

Publisher's PDF, also known as Version of record

General rights

Copyright for the publications made accessible via Ulster University's Research Portal is retained by the author(s) and / or other copyright owners and it is a condition of accessing these publications that users recognise and abide by the legal requirements associated with these rights.

Take down policy

The Research Portal is Ulster University's institutional repository that provides access to Ulster's research outputs. Every effort has been made to ensure that content in the Research Portal does not infringe any person's rights, or applicable UK laws. If you discover content in the Research Portal that you believe breaches copyright or violates any law, please contact pure-support@ulster.ac.uk.

6th
**Multiprofessional Audit
Conference**

**Poster
Abstracts**

**Wednesday, 25th October 2000
The Stormont Hotel**

An Audit of Subcutaneous Syringe Drivers in a Non-Specialist Hospital

**Kathleen Dunne
Kate Sullivan
Angela Garvey
George Kernohan
Angela Diamond
Celine Duffy
Joan Hutchinson**

Altnagelvin Group of Hospitals

DISPLAY 14

The use of syringe drivers as a method of drug delivery to control symptoms in palliative care is a common and accepted practice, but one which has evolved rather than been subject to close multi-professional scrutiny and guideline formation. There is evidence that adverse incidents may arise as a result of syringe driver use (Medical Devices Agency (MDA) 1998), for example, errors in drug calculations, drug stability, equipment failure (including disconnection) and incorrect infusion rates. Inadequate user training, poor servicing of equipment and inadequate documentation and record keeping are all thought to be contributing factors (MDA 1998).

In the hospital where the audit was carried out (a district general hospital without specialist palliative care services), syringe drivers are used to administer drugs to patients with cancer during the palliative phase of illness. The purpose of this clinical audit was to establish the standard of current practice in wards where syringe drivers were being used. A retrospective study of 13 cases of syringe driver use is presented. The results highlight many areas of unregulated practice with regard to setting up, monitoring and maintenance of syringe drivers. The choice of drugs and dosage prescribed, evaluation of treatment responses and review of treatment regimens were also areas of concern. Guidelines for the use of syringe drivers in non-specialist hospitals are put forward.

Correspondence to:

Mrs Kathleen Dunne
Nurse Teacher
North & West In-Service Education Consortium
Altnagelvin Hospital
Londonderry BT47 3SD

Tel: 028 71 345171 Ext 3630
Fax: 028 71 611272