



Effect of varying CRT refresh rate on the measurement of temporal summation.

Zlatkova, M., Redmond, T., Garway-Heath, D., Anderson, R., & Mulholland, P. (2015). Effect of varying CRT refresh rate on the measurement of temporal summation. *Ophthalmic and Physiological Optics*, 35, 582-590. <https://doi.org/10.1111/opo.12227>

[Link to publication record in Ulster University Research Portal](#)

Published in:

Ophthalmic and Physiological Optics

Publication Status:

Published (in print/issue): 24/08/2015

DOI:

[10.1111/opo.12227](https://doi.org/10.1111/opo.12227)

Document Version

Publisher's PDF, also known as Version of record

General rights

Copyright for the publications made accessible via Ulster University's Research Portal is retained by the author(s) and / or other copyright owners and it is a condition of accessing these publications that users recognise and abide by the legal requirements associated with these rights.

Take down policy

The Research Portal is Ulster University's institutional repository that provides access to Ulster's research outputs. Every effort has been made to ensure that content in the Research Portal does not infringe any person's rights, or applicable UK laws. If you discover content in the Research Portal that you believe breaches copyright or violates any law, please contact pure-support@ulster.ac.uk.



TECHNICAL NOTE

Effect of varying CRT refresh rate on the measurement of temporal summation

Pádraig J. Mulholland^{1,2}, Margarita B. Zlatkova², Tony Redmond³, David F. Garway-Heath¹ and Roger S. Anderson^{1,2}

¹National Institute for Health Research (NIHR) Biomedical Research Centre at Moorfields Eye Hospital NHS Foundation Trust and UCL Institute of Ophthalmology, London, UK, ²Vision Science Research Group, School of Biomedical Sciences, University of Ulster, Coleraine, UK, and ³School of Optometry and Vision Sciences, Cardiff University, Cardiff, UK

Citation information: Mulholland PJ, Zlatkova MB, Redmond T, Garway-Heath DF & Anderson RS. Effect of varying CRT refresh rate on the measurement of temporal summation. *Ophthalmic Physiol Opt* 2015; **35**: 582–590. doi: 10.1111/opo.12227

Keywords: cathode-ray-tube, critical duration, refresh rate, stimulus duration, temporal summation

Correspondence: Pádraig J Mulholland
E-mail address: Padraig.Mulholland@moorfields.nhs.uk

Received: 9 January 2015; Accepted: 3 June 2015

Abstract

Purpose: To quantify the effect of cathode-tube-ray (CRT) monitor refresh rate on the measurement of the upper limit of complete temporal summation (critical duration) in the peripheral visual field of healthy observers.

Methods: Contrast thresholds were measured for seven achromatic spot stimuli (diameter 0.48°) of varying duration (nominal values: 10–200 ms) at an eccentricity of 8.8° along the 45°, 135°, 225° and 315° meridians of the visual field in three healthy, psychophysically experienced observers. Stimuli were presented on a CRT display with a refresh rate of 60 and 160 Hz. Contrast thresholds were expressed as contrast energy with stimulus durations being estimated using (1) the sum-of-frames (SOF) method and (2) Bridgeman's method incorporating measurements of phosphor persistence. Estimates of the critical duration were produced using iterative two-phase regression analysis.

Results: With stimulus duration expressed as SOF equivalent the critical duration was, on average, 10.6 ms longer with a refresh rate of 60 Hz (mean 45.7 ms, S.D. 10.1 ms) relative to 160 Hz (35.1 ms, S.D. 7.6 ms). When the Bridgeman method was used, minimal differences (1.8 ms) in critical duration values between the two refresh rates (60 Hz: 33.0 ms, S.D. 9.4 ms; 160 Hz: 31.2 ms, S.D. 7.0 ms) were observed. Identical trends were observed in all three subjects.

Conclusions: Psychophysical measurements of temporal summation are independent of variations in CRT refresh rate when the Bridgeman method, incorporating measured values of phosphor persistence, is used to estimate stimulus duration. This has significant implications for the specification of stimulus duration in psychophysical studies of vision employing conventional display monitors.

Introduction

The value of any visual psychophysical investigation is highly dependent upon the precise control of stimulus presentation parameters. Within the published literature, cathode ray tube (CRT) monitors remain in widespread use for the presentation of psychophysical stimuli despite their lack of production.^{1,2} Image generation in this class of monitor occurs as a result of the activation of individual phosphor particles on the posterior surface of the display screen by an

incoming electron beam. Once activated, each phosphor displays a rapid increase in luminance output followed by an exponential decline in activity until energy emission ceases. Owing to this decay, the re-excitation or *refresh* of each phosphor is required for the desired image to remain on the display screen. The number of occasions each pixel is re-activated in one second determines the refresh rate, with the frame duration being calculated as the reciprocal of this value. Because of the sequential re-activation of each pixel, the temporal delivery of light energy to the eye from

a CRT is intermittent with the rate of flicker being dependent on the refresh rate selected (Figure 1). Despite such periodic temporal output, flicker is not perceived as long as the refresh rate is kept above the critical flicker frequency (CFF). Currently, the majority of studies using a CRT employ refresh rates of 60 Hz (range: 60–200 Hz).²

Although widely used, CRTs and other frame based display monitors (e.g. organic light emitting diodes) are not ideal for the psychophysical examination of vision. Their primary limitation relates to temporal display artefacts resulting from the image generation process.^{1–4} Specifically, CRT monitors are unable to properly replicate stimuli with square wave temporal profiles due to phosphor decay and reactivation.^{1,2} This limitation introduces two specific issues for the use of CRT monitors in vision science – (1) the generation of neural artefacts that can potentially influence the results of any psychophysical experiment and (2) difficulties in meaningfully specifying the duration of stimuli.

Within the literature, the potential for neural artefacts arising from the pulsed nature of CRT output has been widely discussed.³ Gawne and Woods,⁵ in a neurophysiology study, report that pulsed stimuli with a gradual offset, such as those generated on a CRT, do not produce responses in neurons within cortical area V1 that are comparable to those gained when using a true square wave stimulus of equal nominal duration. Zele and Vingrys⁶ also propose neural artefacts to occur at the level of the retina due to the formation of high-frequency noise, this effect being amplified when lower refresh rates are used. Shady et al.⁷ in a psychophysical study reported that adaptation to flicker could affect visual thresholds even if the flicker

frequency is above the CFF. They point out that, even if flicker is not perceived, low CRT refresh rates can influence visual sensitivity, advising that high refresh rates should be used where possible to avoid artefactual deficits in visual sensitivity. Despite evidence that variations in refresh rate can influence some neural processing in the visual pathway,^{5,6} it is unknown if such changes can also influence the integration of light photons over time (temporal summation), and specifically the critical duration (Figure 2).

The inability of a CRT to reproduce stimuli with a square wave temporal profile also creates problems when attempting to specify the duration of psychophysical stimuli. Square wave stimuli generated on a CRT characteristically have a rapid onset, a variable temporal profile and display a tapered offset due to phosphor decay in the final frame of presentation (Figure 1). The most commonly used method to estimate presentation duration is the sum-of-frames (SOF) calculation. This method calculates presentation duration on the basis that stimuli may only be integers of frames when generated on a CRT display and assumes contiguous light output over the duration of the stimulus with no allowance being made for phosphor activation and decay within each frame.^{2,3} The SOF method, although convenient, can lead to significant over-estimation of stimulus duration, this being amplified for single frame presentations where the period of phosphor activation is shorter than the frame duration (Figure 3a). In response to such limitations, Bridgeman¹ proposed that stimulus duration be measured from the point of phosphor activation in the first frame to the temporal limit of phosphor activity in the final frame of presentation. Although theoretically superior




Figure 1. Comparison of the temporal profile of luminance output for a 100 ms stimulus as measured on a CRT display running at 60 Hz (upper panel) and 160 Hz (lower panel).




Figure 2. Schematic temporal summation function. For short duration stimuli there is complete summation (grey shaded area) and the data may be fit with a line of slope zero for calculated energy data up to the critical duration (blue arrow). Beyond the critical duration incomplete summation is exhibited.




Figure 3. (a) Schematic detailing how phosphor persistence/decay time (p) was calculated from phosphor activity plots. Dashed lines indicate the start of the frame, decay time and end of frame (refresh 160 Hz). Stimulus duration as specified using the SOF and modified Bridgeman methods are also listed for reference. (b,c) Phosphor decay times measured for the P45 phosphor at a range of (b) luminance and (c) energy output levels (E_{stim} for single frame presentation of one pixel area) for a CRT running at 60 (blue circles) and 160 Hz (red squares).

to the SOF method, knowledge of the phosphor decay time is required, it also being unknown if this value is affected by luminance output. Furthermore, as the SOF method is used almost exclusively in the literature to specify stimulus duration on display monitors² it is currently unknown what effect, if any, using the Bridgeman method will have on the results of a psychophysical study of temporal vision.

In this study, we sought to investigate the effect of varying refresh rate on measurements of temporal summation for an achromatic spot stimulus, generated on a CRT display under photopic conditions. The purpose of this investigation was twofold. Primarily we wished to determine if the selection of a low (60 Hz) or a high refresh rate (160 Hz), in those studies employing frame based display monitors, impacts upon the measurement of temporal summation in healthy observers. We also wanted to determine the effect of specifying stimulus duration using the SOF and Bridgeman methods on the perceived trends within a psychophysical study of temporal vision.

Methods

Subjects

Three healthy volunteers, aged 25, 31 and 48 years with normal or corrected-to-normal vision, were included in this study. These included two of the authors (PJM and RSA) and one naive observer (NCS). Corrected visual acuity was $-0.10 \log\text{MAR}$ (Snellen 6/5 or 20/17) with a refractive error within ± 3 dioptres (D) and 0.50 D astigmatism in all subjects. Each subject's right eye was examined with a natural pupil (5–7 mm diameter). Ethical approval was gained from the London-Central National Research Ethics

Service committee and the research protocol adhered to the tenets of the Declaration of Helsinki.

Apparatus & stimuli

Stimuli were presented on a γ -corrected 21" Phillips FIMI MGD-403 achromatic CRT monitor (www.ampronix.com) with a pixel resolution of 500×720 . Two refresh rates of 60 Hz and 160 Hz were used. A ViSaGe MKII Visual Stimulus Generator (ViSaGe, Cambridge Research Systems, www.crsrtd.com) and Cambridge Research Systems toolbox (v.1.27) for MATLAB (version R2011a, www.mathworks.com) were used to generate stimuli. Chromaticity co-ordinates of stimuli, background and central fixation cross were $x = 0.258$ and $y = 0.257$ as measured using a colorimeter (ColorCal MKII, Cambridge Research Systems, www.crsrtd.com). For all trials, circular stimuli of diameter 0.48° (equivalent to a Goldmann size III clinical perimetric stimulus) were presented on a background of 10 cd m^{-2} . The CRT display was viewed from a distance of 60 cm with subjects placing their head in a chin-rest during examinations. All subjects were optically corrected for the test distance using full aperture trial lenses.

The temporal profile of luminance output from the CRT display, in addition to the refresh rate of the display, was measured using an Optical Transient Recorder 3 (OTR-3, www.display-messtechnik.de/) configured for unipolar output. To permit full measurement across the complete range of possible contrast levels, a gain or amplitude setting of S3 and variable voltage range between 1 and 10 V were employed, together with a receiver aperture of 3 mm. Prior to any data collection measurements were performed in the

absence of light (five measurements of 1 s duration) to account for any noise within the OTR-3 device. All other OTR-3 recordings were normalized using the mean amplitude value in dark conditions as a baseline, any excursion beyond this point representing light output.

Measurement of phosphor activity

Circular spot stimuli of diameter 42 mm and single frame duration were presented at one of the test locations (8.8° eccentricity along 45° meridian). The OTR-3 was positioned so that the centre of the receiver aperture was perpendicular to, and coincidental with, the centre of the test stimulus. Measurements were repeated for all contrast levels in an otherwise dark room.

Estimated stimulus contrast energy

To estimate contrast energy (ΔE) from luminance values, we assumed the CRT luminance output to be a square wave, with Equation 1 then used to estimate ΔE for stimuli of varying duration. The value L corresponds to the luminance measurement collected by the ColorCal II, L_b the background luminance, f the stimulus duration (expressed as number of frames) and r the refresh rate.

$$\Delta E = \left(\frac{f}{r}\right)(L - L_b) \quad (1)$$

Calculation of stimulus duration

Stimulus duration was calculated using both the SOF (t_{sof}) and Bridgeman methods (t_{bn}). Equation 2 was used to calculate SOF durations in ms where f is the number of frames within the stimulus and r the refresh rate.

$$t_{\text{sof}} = f \left(\frac{1000}{r}\right) \quad (2)$$

Stimulus duration was also estimated from the point of phosphor activation in the first frame to the temporal limit of activity in the final frame of presentation (Bridgeman method, Equation 3).

$$t_{\text{bn}} = \left[(f - 1) \left(\frac{1000}{r}\right) \right] + p \quad (3)$$

Bridgeman¹ suggests that a constant value for phosphor persistence (p), or decay time, be incorporated in the calcu-

lation. Unfortunately the percentage decay to which p should be measured (i.e. temporal limit of phosphor activity), together with the point above zero output that defines the start of phosphor activity, was not specified. For the purposes of this study we specified both the start and end of phosphor activity within a frame to be 10% above baseline (Figure 3a). When plotted as a function of luminance output (Figure 3b) and energy values (Figure 3c, see supplementary information for calculation of output energy from OTR-3 measurements), no change in p was observed (60 Hz: $r^2 = 0.11$; 160 Hz: $r^2 = 0.10$, both $p > 0.05$ for r^2 values). In view of this p was calculated as the mean of all measurements collected (1.8 ms).

Psychophysical procedure

Two subjects (RSA and NCS) underwent one complete examination for each refresh rate. The experiment was performed twice at each refresh rate (in a random order) for subject PJM. In each experiment, contrast thresholds were measured for achromatic spots (0.48°) of varying nominal duration (10–200 ms) at 8.8° eccentricity in the visual field along the 45°, 135°, 225° and 315° meridians, in an interleaved fashion. A yes/no response paradigm was employed with a 1/1 staircase that terminated after six reversals. Threshold luminance was calculated as the mean of the final four reversal values at each test location. Subjects were instructed to fixate a central cross target and press a response button if a stimulus was seen. Reliability was assessed with blank presentations (false positive catch trials) that accounted for approximately 30% of all presentations. The session was halted and repeated if the false positive rate exceeded 20%. Prior to data collection, subjects were given one or more practice sessions, until it was clear that they fully understood the task.

Data analysis

Temporal summation functions were generated with stimulus durations calculated as SOF (Equation 2) and Bridgeman equivalents (Equation 3). Each summation function, expressed as $\log \Delta E$ vs \log stimulus duration, was constructed for each subject using thresholds (mean across all test locations and test runs) for stimuli of different durations. Two-phase regression analysis⁸ was used to estimate the critical duration from the temporal summation curves. As part of this analysis, the slope of the first line was constrained to 0 in accordance with Bloch's law (complete temporal summation). The slope and intercept of the second line, along with the point at which the two component lines met (breakpoint), were free to vary. The critical duration was estimated, following multiple iterations (maximum 1000), as the breakpoint in the function.

Results

When stimulus duration was expressed as SOF, equivalent critical duration values were greater for stimuli presented with the lower refresh rate of 60 Hz (mean 45.7 ± 10.1 ms) compared with 160 Hz (mean 35.1 ± 7.6 ms). This trend was seen for all subjects with a mean difference of 10.6 ± 2.8 ms (Figure 4, upper panel). This difference was statistically significant when examined with a paired *t*-test ($p = 0.02$). When the Bridgeman method was used to estimate stimulus duration, minimal differences (mean 1.8 ± 2.8 ms) in critical duration were observed with refresh rate (Figure 4, lower panel). These differences were not statistically significant ($p = 0.43$ in a paired *t*-test). Mean critical duration values for the 60 Hz and 160 Hz frame rates were 33.0 ± 9.4 ms and 31.2 ± 7.0 ms, respectively. Unsurprisingly, the critical duration values were shorter when stimulus duration was expressed using the Bridgeman method compared to the SOF method. If the data collected using the 60 Hz display are considered, critical duration values are on average 12.7 ms shorter using the Bridgeman durations compared to the SOF equivalent. For the same method of threshold expression,

the discrepancy was much smaller (3.9 ms) for the 160 Hz data set.

The discrepancy in stimulus durations when expressed as SOF and Bridgeman equivalents (mean values across all locations and subjects in study) for each nominal stimulus duration (i.e. those specified in experimental code) may be seen more clearly in Figure 5. The SOF method consistently yields higher estimates of stimulus duration across the range of stimuli presented in this study. This discrepancy is greatest for the lower refresh rate of 60 Hz. It may also be seen that for stimuli of single frame duration the SOF method can introduce particularly large errors, these inaccuracies being greatest for displays running with a low refresh rate.

Discussion

Temporal summation with variations in refresh rate

The present study shows that the critical duration of temporal summation for a perceptually single achromatic spot stimulus is independent of CRT refresh rate when the Bridgeman method, incorporating measured values of




Figure 4. Temporal summation functions for threshold data expressed as contrast energy values for individual subjects and stimulus durations specified as SOF equivalent (upper panel) and Bridgeman values (lower panel). Error bars included represent the standard error of the mean (S.E.M.). The breakpoint in each function (dashed line) indicates the critical duration.




Figure 5. Comparison of stimulus duration as estimated using the SOF (squares) and Bridgeman methods (circles) for stimuli generated on display with 60 (blue) and 160 Hz frame rates (red). Nominal stimulus durations represent the duration specified in the experimental code. The number of constituent frames in each stimulus is included as a label on the Bridgeman data points.

phosphor persistence, is used to estimate stimulus duration. Although no previous experiment has investigated the temporal summation of a CRT signal, a variety of studies have explored the summation of pairs of incremental stimuli presented with varying temporal separations. A finding common to these studies is the complete summation of energy for temporally double-pulsed spot stimuli when presented with short inter-stimulus intervals up to a critical duration.^{9–12} After this partial summation is observed until a point is reached at a separation of approximately 60 ms where cancellation or inhibition is seen to occur.^{11,12} Such trends have been attributed to the presence of bi-phasic temporal filters in the visual system.¹¹

If the response to the temporally modulated stimulus at threshold is considered to be mediated by a linear filter (see Watson¹³ for a review), it can be shown that the visual thresholds within the critical duration should not change with the refresh rate of the monitor. It has been proposed that the response of the visual system will be constant if the product of the amplitude spectra of the stimulus and amplitude response of the linear filter is equal within the critical duration.¹³ Assuming that the amplitude response remains constant within each subject under the conditions of this experiment (i.e. identical background luminance, stimuli, etc.), it can be seen from the amplitude spectra (Figure 6, lower panel) that the peak amplitude (1st harmonic) is identical for the 60 Hz and 160 Hz stimuli.

Treating the visual system as a linear filter with a certain amplitude response with a maximum at 7–8 Hz and a cut off frequency at about 40 Hz,¹⁴ it can be seen that increasing the refresh rate of the display should not significantly influence the product (convolution) of the amplitude spectra of the stimulus and the amplitude response of the linear filter in the range of maximal response, with the result that visual thresholds and thus the critical duration will remain invariant of refresh rate.

One study has, however, challenged the notion that the visual system may completely sum stimuli presented on a CRT display with a low refresh rate. Using a large stimulus diameter (17°) and high retinal illuminance (700 trolands), Rashbass¹² found summation to be incomplete when the time interval between two successive incremental pulses was 8 ms, this being noticeably shorter than the critical duration (16 ms) found under identical test conditions for a single stimulus of equal total duration and area. One possible explanation for the discrepancy between the work of Rashbass¹² and the results of this study is the experimental conditions used. The temporal summation of single stimuli is known to be influenced by a number of factors including stimulus area^{15,16} and background adapting luminance,¹⁷ with a shorter critical duration at higher adapting luminance and larger stimulus size. In a similar fashion, the summation of stimuli composed of multiple incremental pulses is affected by factors relating to both the stimulus and environment.^{11,13} It is thus likely that the relatively smaller stimulus (0.48°) and lower background luminance (10 cd m⁻²) used in this investigation would lead to a longer critical duration in a temporal double-pulse experiment and, as a result, no difference in the critical duration with refresh rate.

Specifying stimulus duration

The inherent difficulty in estimating the duration of a stimulus presented on any display monitor has been widely reported in published literature.^{1–3,18} In agreement with previous work, the SOF method, as applied in this study, appears to overestimate durations for stimuli with a small number of constituent frames.^{1,4,18} Significantly, these disparities appear to be greater when the lower refresh rate of 60 Hz was selected (Figure 5). Considering the example of a nominal 10 ms stimulus reproduced on a display with a 60 Hz refresh rate, the SOF estimation of duration (one frame, 16.7 ms) is 828% greater than the Bridgeman equivalent (1.8 ms) for the group of subjects in this study. For the same stimulus generated on a display running at 160 Hz (two frames, $t_{\text{sof}} = 12.5$ ms, $t_{\text{bn}} = 8.1$ ms), the discrepancy is smaller (54%). These differences and their relative effect on psychophysical thresholds are, however, partly dependent upon the type of phosphor used. Di Lollo




Figure 6. Schematic temporal profile (upper panel) of threshold stimuli of duration shorter than the critical duration (12 & 24 ms, equal total energy) generated with temporal frequencies of 1 Hz (leftmost plot, black lines), 60 Hz (centre plot, blue lines) and 160 Hz (rightmost plot, red lines) along with corresponding amplitude spectra (lower panel). The 1 Hz frequency is included for illustration only as reference to a true square wave stimulus.

*et al.*¹⁹ found the persistence of the P31 phosphor to be visible several hundred milliseconds after presumed stimulus offset in dark-adapted conditions and also in the presence of a 'veiling glare' (achieved with two lamps with an output attenuated to 0.33 cd m^{-2}). This effect was amplified for displays using phosphors of high persistence and stimuli of high luminance.

A number of authors have questioned the value in accurately specifying the duration of stimuli when shorter than the critical duration.¹⁸ It is well established that spatial and temporal resolution decrease with increasing levels of summation,²⁰ thus if a stimulus is shorter than the critical duration, the visual system will only differentiate on the basis of luminous flux and not duration. The results of this study present a strong argument against this view. When examining the temporal aspects of vision, such as summation, it is clear that small discrepancies in stated duration can induce large deviations from the true trends in a given data set. Elze⁴ in an examination of simulated frequency-of-seeing data found the maximum likelihood method used to generate each psychometric function to be influenced by the

method used to estimate stimulus duration. This difference was attributed to lack of assumed proportionality of the SOF method compared with the Bridgeman calculation. More simply, a stimulus composed of two frames is assumed by the SOF to be double the duration of a single frame presentation. This is not the case when duration is specified as a Bridgeman equivalent. In a similar fashion the results of the iterative two-phase regression analysis used to estimate the critical duration in this study was also influenced by the method chosen to estimate stimulus duration.

In this study, Bridgeman's method was exclusively applied to estimate the duration of stimuli generated on a CRT display. The use of this calculation may, however, be also extended to describing the duration of stimuli produced on other display types such as organic light-emitting diode (OLED) monitors whose pulsed output resembles that of a CRT.²¹ Although the temporal output of OLED monitors varies from that of a CRT (i.e. a more rapid decay to 0% of peak output within a frame) there appears to be a period within each frame where no energy output takes

place. Ito *et al.*²¹ demonstrated that a stimulus alternating in RGB values from (255, 255, 255) to (192, 192, 192) with each frame refresh on a Sony PVM-2541 OLED monitor (refresh rate 60 Hz) led to periods of light emission (~7.5 ms) followed by intervals (~9.2 ms) where no light output was detected. Considering this evidence it is also likely that very short duration stimuli produced on OLED displays might also suffer from over-estimations of stimulus duration should the SOF method be used. In this situation the Bridgeman method incorporating persistence (*p*) values equal to the period of light emission in a single frame could be applied to improve the accuracy of any estimates of stimulus duration.

Refresh rate selection

The issue of temporal presentation artefacts associated with display monitors, together with methods for their reduction, has been widely discussed within the psychophysical literature. Specifically, temporal variations in luminance output secondary to phosphor decay in CRT displays have been highlighted as a drawback when attempting to accurately estimate the duration of stimuli presented and also replicate stimuli with square wave temporal profiles.^{1–3} To partially alleviate such issues, it has been suggested that a high refresh rate should be employed. This assertion appears to have been made without regard to how the visual system sums the temporal output from a CRT display or whether varying refresh rate impacts upon psychophysical thresholds. It is clear from the results of this study that the upper limit of complete temporal summation remains constant *in contrast energy terms* despite variations in the nature of energy delivery resulting from changes to refresh rate. Despite potentially influencing the activity of retinal ganglion cells⁶ and cortical neurons in area V1,⁷ low refresh rates do not appear to impact upon the investigation of temporal vision provided output from the CRT display is accurately characterised in terms of both energy and duration using appropriate metrics.

A wide range of CRT refresh rates have been selected for use in both the clinical and basic psychophysical examination of vision, in order to reduce neural artifacts,^{6,22} improve temporal resolution,³ reduce flicker perception at high background luminance¹ and also reduce the effects of adaptation to invisible flicker on visual sensitivity.⁷ In this study, we have demonstrated that the selection of refresh rate may also have an effect on the ability to accurately specify stimulus duration, thus leading to secondary and, most importantly, artificial variations when investigating temporal visual processing. Interestingly, the difference between critical duration values estimated using SOF stimulus durations, compared with the more accurate Bridgeman durations, was smallest when a high refresh rate was

used. This finding may be due to an improved correspondence between the temporal profile of a stimulus produced with a high refresh rate and the contiguous energy output assumed by the SOF method of classifying stimulus duration on a CRT display. As the measurement of energy output, or indeed phosphor decay time, may not be practicable in all situations, it is strongly advisable that, when using the SOF method, a high refresh rate be used where possible to reduce any disparities between the *real* and estimated stimulus durations.

Conclusions

CRT displays continue to offer psychophysicists the ability to present a wide variety of accurately calibrated visual stimuli. The capability of the visual system to sum energy delivered over a given temporal window appears to be independent of duty cycle changes, secondary to variations in refresh rate, for an achromatic stimulus of 0.48° diameter. It is clear from the results of this study that the quantification of CRT output, specifically presentation duration, can greatly impact upon the investigation of temporal vision using this class of display monitor. The use of accurate metrics that make reference to the *real* temporal profile of monitor output partially alleviate such issues and have the potential to serve as universal metrics through which data collected using varying CRT refresh rates, or indeed of different monitor types, may be accurately, and more importantly, validly compared.

Disclosure

Four of the authors (DFG-H, RSA, TR and PJM) have financial disclosures that may be of relevance to this work. These are: Research support: Carl Zeiss Meditec (DFG-H), Heidelberg Engineering (TR, DFG-H, RSA); Honoraria: Carl Zeiss Meditec (DFG-H), Heidelberg Engineering (DFG-H, RSA, PJM); Co-inventor: Moorfields Motion Detection Test (DFG-H).

Acknowledgements

This work was supported by a PhD studentship from the Department for Employment and Learning, Northern Ireland (PJM), and in part by the National Institute for Health Research (NIHR) Biomedical Research Centre based at Moorfields Eye Hospital NHS Foundation Trust and UCL Institute of Ophthalmology (PJM, DFG-H). DFG-H's chair at UCL is supported by funding from the International Glaucoma Association. The views expressed are those of the authors and not necessarily those of the NHS, the NIHR or the Department of Health. The article was presented in part as a poster at the ACM Symposium on

Applied Perception (SAP) annual meeting, Dublin, Ireland in 2013. Finally, we would like to thank the anonymous reviewers for their helpful comments on this manuscript.

References

1. Bridgeman B. Durations of stimuli displayed on video display terminals: $(n-1)/f + \text{persistence}$. *Psychol Sci* 1998; 9: 232–233.
2. Elze T. Misspecifications of stimulus presentation durations in experimental psychology: a systematic review of the psychophysics literature. *PLoS One* 2010; 5: 1–7.
3. Bach M, Meigen T & Strasburger H. Raster-scan cathode-ray tubes for vision research—limits of resolution in space, time and intensity, and some solutions. *Spat Vis* 1997; 10: 403–414.
4. Elze T. Achieving precise display timing in visual neuroscience experiments. *J Neurosci Methods* 2010; 191: 171–179.
5. Gawne TJ & Woods JM. Video-rate and continuous visual stimuli do not produce equivalent response timings in visual cortical neurons. *Vis Neurosci* 2003; 20: 495–500.
6. Zele AJ & Vingrys AJ. Cathode-ray-tube monitor artefacts in neurophysiology. *J Neurosci Methods* 2005; 141: 1–7.
7. Shady S, MacLeod DI & Fisher HS. Adaptation from invisible flicker. *Proc Natl Acad Sci USA* 2004; 101: 5170–5173.
8. Seber GAF & Wild CJ. *Nonlinear Regression*, John Wiley & Sons: New York, 1989.
9. Granit R & Davis W. Comparative studies on the peripheral and central retina IV: temporal summation of subliminal visual stimuli and the time course of the excitatory after-effect. *Am J Physiol* 1931; 98: 644–653.
10. Davy E. The intensity-time relation for multiple flashes of light in the peripheral retina. *J Opt Soc Am* 1952; 42: 937–941.
11. Ikeda M. Temporal summation of positive and negative flashes in the visual system. *J Opt Soc Am* 1965; 55: 1527–1533.
12. Rashbass C. The visibility of transient changes of luminance. *J Physiol* 1970; 210: 165–186.
13. Watson AB. Temporal sensitivity. In: *Handbook of Perception and Human Performance* (Boff KR, Kaufman L & Thomas JP, editors), Wiley: New York; Chichester, 1986; Chapter 6; pp. 1–43.
14. Robson JG. Spatial and temporal contrast sensitivity functions of the visual system. *J Opt Soc Am* 1966; 56: 1141–1142.
15. Karn H. Area and the intensity-time relation in the fovea. *J Gen Psychol* 1936; 14: 360–369.
16. Barlow HB. Temporal and spatial summation in human vision at different background intensities. *J Physiol (Lond)* 1958; 141: 337–350.
17. Graham CH & Kemp EH. Brightness discrimination as a function of the duration of the increment in intensity. *J Gen Physiol* 1938; 21: 635–650.
18. Krantz JH. Tell me, what did you see? The stimulus on computers. *Behav Res Methods Instrum Comput* 2000; 32: 221–229.
19. Di Lollo V, Seiffert AE, Burchett G, Rabeeh R & Ruman TA. Phosphor persistence of oscilloscopic displays: a comparison of four phosphors. *Spat Vis* 1997; 10: 353–360.
20. Barlow HB & Mollon JD. Psychophysical measurements of visual performance. In: *The Senses* (Barlow HB & Mollon JD, editor), Cambridge University Press: Cambridge, 1982; pp. 114–132.
21. Ito H, Ogawa M & Sunaga S. Evaluation of an organic light-emitting diode display for precise visual stimulation. *J Vis* 2013; 13: 1–21.
22. Zele AJ, O’Loughlin RK, Guymer RH & Vingrys AJ. Disclosing disease mechanisms with a spatio-temporal summation paradigm. *Graefes Arch Clin Exp Ophthalmol* 2006; 244: 425–432.