



New Developments in Smart Bandage Technologies for Wound Diagnostics

McLister, A., McHugh, J., Cundell, J., & Davis, J. (2016). New Developments in Smart Bandage Technologies for Wound Diagnostics. *Advanced Materials*, 28(27), 5732-5737. <https://doi.org/10.1002/adma.201504829>

[Link to publication record in Ulster University Research Portal](#)

Published in:
Advanced Materials

Publication Status:
Published (in print/issue): 20/07/2016

DOI:
[10.1002/adma.201504829](https://doi.org/10.1002/adma.201504829)

Document Version
Author Accepted version

General rights
Copyright for the publications made accessible via Ulster University's Research Portal is retained by the author(s) and / or other copyright owners and it is a condition of accessing these publications that users recognise and abide by the legal requirements associated with these rights.

Take down policy
The Research Portal is Ulster University's institutional repository that provides access to Ulster's research outputs. Every effort has been made to ensure that content in the Research Portal does not infringe any person's rights, or applicable UK laws. If you discover content in the Research Portal that you believe breaches copyright or violates any law, please contact pure-support@ulster.ac.uk.

DOI: 10.1002/ ((please add manuscript number))

Article type: ((Research News))

New Developments in Smart Bandage Technologies for Wound Diagnostics

Anna McLister¹, Jolene Phair¹, Jill Cundell², James Davis^{1}*

¹School of Engineering, ²School of Health Sciences, University of Ulster, Jordanstown, Northern Ireland, BT37 0QB, UK

E-mail: james.davis@ulster.ac.uk

Prof J. Davis

School of Engineering, University of Ulster, Jordanstown, Northern Ireland, BT37 0QB, UK

Keywords: Electrode, Sensors, Wound, Healing, Infection

Abstract: The pH of wound fluid has long been recognised as an important diagnostic for assessing wound condition but as yet there are few technological options available to the clinician. The availability of sensors that could measure wound pH, either in the clinic or in home could significantly improve clinical outcome – particularly in the early identification of complications such as infection. This review identifies new material designs and electrochemical research strategies that are being targeted at wound diagnostics and provides a critical overview of emerging research that could be pivotal in setting the direction for future devices.

1. Introduction

Small cuts and scratches are part and parcel of everyday life and most people, providing the trauma to the skin is relatively minor, will cede responsibility for repair of the wound to the myriad of biochemical processes that govern wound healing. In effect, beyond the simple cleansing of the wound, the body will be expected to act to re-establish the integrity of the skin barrier on its own accord with only the briefest oversight from the patient. This is not however always the case and the healing processes that would normally regulate tissue regeneration can become slowed or stalled^[1-5]. The latter can arise as a consequence of

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19
20
21
22
23
24
25
26
27
28
29
30
31
32
33
34
35
36
37
38
39
40
41
42
43
44
45
46
47
48
49
50
51
52
53
54
55
56
57
58
59
60
61
62
63
64
65

numerous factors such as infection, compromised nutritional status and poor circulatory supply. The development of chronic wounds has long been a major concern for healthcare providers and especially so for those involved in the management of diabetic patients where ulceration is a common and increasingly problematic complication^[1-3]. While advances in sensing technologies have revolutionised the daily management of diabetes through enabling patients to monitor their blood glucose levels, there is an emerging opportunity for these same methods to provide valuable diagnostic insights into the healing process and therein proffer the possibility of markedly aiding the treatment of chronic wounds.

Chronic wounds are widely regarded as a silent epidemic that affects a significant proportion of the populace and pose a major and, indeed, an ever increasing threat to public health^[1] and while, in recent years, there has been a significant increase in the number and variety of wound dressings^[4], the management of the condition continues to be problematic. In chronic wounds, typified by diabetic foot, venous leg or pressure ulcers, the healing processes are stalled resulting in a wound that can persist for months if not years leaving the patient susceptible to further complications and life threatening events^[1-3]. It is estimated that some 6.5 million people in the US suffer from a chronic wound of one sort or another giving rise to an annual healthcare bill of some \$25 billion ^[6]. Admittedly, the US health system is far from typical, but figures of a similar magnitude can be found in most developed countries. In the UK, some 650,000 patients suffer from some form of chronic wound with published estimates suggesting that the combined cost to the NHS for their treatment reaches approximately £3 billion per year^[1]. Only 5% of the latter is attributed to materials cost with the vast majority – estimated to be around 80% - being staff-related. In a recent audit almost 42% of leg/foot ulcers had not healed in the previous six months and 28% had remained unhealed for a year or longer which, in all cases, requires continuous outpatient management, hospital consultation and treatment^[1]. It must also be noted that the loss of productivity for afflicted individuals and the families that care for them and their diminished quality of life are

1 simply immeasurable and will far outweigh the fiscal concerns of the state. Sadly, the burden
2 of treatment is expected to increase dramatically in the future with ever-rising healthcare costs,
3 an aging population and dramatic increases in diabetes and obesity (critical comorbidities).
4 Similarly, burn injuries are another significant cause for concern for clinicians involved in
5 complex wound care (and NHS budget holders) with over a quarter of a million cases
6 presenting to primary care teams and a further 175,000 passing directly to A&E departments
7 annually^[7].
8
9

10
11
12
13
14
15
16
17 The important point to note when considering these statistics is that, in the vast
18 majority of cases, the management and treatment of the wounds are largely conducted within
19 the community. There is a dearth of diagnostic options available for the routine, point of care
20 assessment of wound condition with the patient and/or healthcare professional relying on the
21 recognition of subtle local indicators or non-specific general signs (such as loss of appetite,
22 malaise, or deterioration of glycaemic control in diabetic patients). Given the great variation
23 in wound type, ambiguity is an ever-present hazard. The longer a wound takes to heal
24 (irrespective of type/origin), the greater the propensity for complications to arise and, where
25 there is ambiguity or a lack of vigilance in assessing wound condition, delays in seeking
26 medical attention can all too often lead to an irreparable deterioration in the wound condition
27 and further compromise the health of the patient leading to hospitalisation with either limb or
28 life-threatening consequences.
29
30
31
32
33
34
35
36
37
38
39
40
41
42
43
44
45

46 As the majority of chronic wounds are treated within the community, the ultimate aim
47 is for the development of intelligent, decentralised, wound care technologies that can monitor
48 the condition of the wound - reporting directly to the patient and/or healthcare practitioner and,
49 where appropriate, able to act autonomously to facilitate the healing processes or minimise
50 complications such as inflammation or infection. It is easy to envisage therefore that the
51 development of new technologies that can reduce treatment times and minimise complications
52 will have major impact and deliver substantial cost benefits and vastly improve patient
53
54
55
56
57
58
59
60
61
62
63
64
65

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19
20
21
22
23
24
25
26
27
28
29
30
31
32
33
34
35
36
37
38
39
40
41
42
43
44
45
46
47
48
49
50
51
52
53
54
55
56
57
58
59
60
61
62
63
64
65

outcome. One of the key findings in the recent World Union of Wound Healing Societies' report: 'Diagnostics and Wounds: A Consensus Document' was that "diagnostic tools need to be moved into the clinic or the patient's home to ensure optimal care is provided for patients with wounds"^[8]. At present, there are no available technologies to address this recommendation.

In this report, a spotlight is trained on the core issues of chronic and complex wounds which, can be considered to centre on four interdependent themes. These are based on monitoring healing, understanding the dynamics of healing and the material interactions, treatment interventions and the electronics for facilitating the intelligence needed for establishing control over the other three. An indication of the convergence of specialisms that are necessary for the development of a smart dressing are indicated in Figure 1. The difference between a bandage and dressing should be noted as the latter is usually in contact with the wound while the former is responsible for covering and fixing the position of the dressing. The advent of new approaches to monitoring wound status condition can lead to a blurring of these definitions as the integration of functional dressing materials with communication technologies embedded with the bandage requires a much more holistic design strategy.

An extensive range of materials that have been developed for wound dressings in recent years and there has been a gradual evolution from the purely passive, biocompatible absorbant polymers to more functional materials that aim to actively encourage the healing processes and minimise the risk of infection. These tend to incorporate components such as metalloproteinase inhibitors and/or antimicrobial agents such as iodine or silver^[4]. The various strategies available have been critically reviewed by Moura et al, (2013) and it is clear that, despite considerable advances in clinical efficacy, healthcare staff remain effectively blind to the progress of the healing processes until the bandage is removed ^[4,7,8]. A much

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19
20
21
22
23
24
25
26
27
28
29
30
31
32
33
34
35
36
37
38
39
40
41
42
43
44
45
46
47
48
49
50
51
52
53
54
55
56
57
58
59
60
61
62
63
64
65

better solution would be to have a dressing that can proactively monitor the wound and provide an insight into the wound dynamics. The latter is a considerable challenge in terms of designing smart materials that can function as the interface between the wound and the electronics and although it is a highly active area, it is one fraught with problems: biocompatibility, selectivity, sensitivity and sensor lifespan being only a few. The aims of the present report are to provide a critical insight in to the material approaches being taken at present in the design of smart dressings and to appraise their translation to the clinic.

2. Wound Healing

Wound healing can be divided into four overlapping phases: coagulation, inflammation, migration-proliferation and remodelling^[5,9]. Coagulation is needed for wound protection and hemostasis and occurs in the very early stages of an injury whereby inflammatory cells are sent to the site of the wound and a protective fibrin plug formed. As the inflammatory phase subsides, wound contraction is initiated. Matrix proteins provide substrates for cell movement, which facilitates the change in cell behaviour and structures that eventually returns the integrity of the tissue^[5,9]. In contrast, chronic wounds seldom progress beyond the inflammatory stage and can be defined more specifically as “a wound that has not shown a 20%-40% reduction in area after 2-4 weeks.”^[2,9]. The feedback mechanisms that would normally end the Inflammatory Stage are short-circuited and are the main reasons for impaired wound healing^[9]. Although Infection is the primary concern in the management of chronic wounds, there is a tendency for individuals to seek treatment only once gross symptoms appear (yellow exudate and red inflammation) by which time bacterial colonisation will have progressed to the point where more substantial intervention is required. Early identification of infection is imperative in minimising the need for hospitalisation and preventing limb threatening events^[10] and it is here that electrochemical sensing systems could have a major impact on clinical practice.

3.0 Wound Fluid

The fluid bathing the wound originates from a variety of sources and comprises a highly heterogeneous mixture and its composition is assumed to reflect the clinical condition of the wound at the time of sampling ^[11,12]. The composition of the fluid will vary considerably over the course of the healing process as a consequence of changes in the microenvironment as cell migration and tissue remodeling progresses and knowing what stage the wound is presently at is one of the driving forces behind the development of wound monitoring technologies. Underlying conditions (i.e. diabetes, neuropathy) can also influence the nature of the fluid. Exogenous factors such as bacterial load, treatments applied (either systemic or topical) and the nature of the wound dressing will also contribute to and affect the milieu and again, the ability to assess the changes in the wound environment through simple diagnostics test could revolutionise treatment.

In chronic ulcers, wound fluid is normally defined as an exudate with a high viscosity and protein content that exceeds 30 mg/mL^[11]. Temporal issues are critical when considering the fluid under examination as it has been found by Zillmer at al. and colleagues that immediately following surgery, there can be significant contamination from material being transported directly from the blood stream – accounting for some 30% of the proteinaceous material found with the wound fluid^[13]. It is important to consider the etiologies involved in the samples under consideration (ie comparing acute to chronic wounds) whereby the concentrations of target biomarkers in one particular sample may not correlate to the actual bioactivity in another.

4.0 Potential Biomarkers

Broadbent and co-workers have detailed over 150 proteins and small molecule metabolites whose concentrations are influenced by the cellular/tissue remodeling and

1 associated inflammatory responses^[14]. The up regulation of proteases and growth factor
2 dysregulation have been identified as offering considerable therapeutic and prognostic value.
3
4 At present, there is a tremendous effort to elucidate the significance of each of these players
5
6 and to identify those most likely to be diagnostically useful in the clinical management of
7
8 chronic wounds. Monitoring markers that can provide key warnings towards the
9
10 complications, particularly infection, is the critical challenge that faces the development of the
11
12 next generation of smart dressings. Clinical investigations of wound fluids, exudates and
13
14 tissues have provided considerable insights and there have been a number of attempts to
15
16 classify the main protagonists involved in the healing progression of the wound. As
17
18 mentioned previously, the ability to control the activation of such species has been shown to
19
20 aid the healing process but such interventions clearly require the provision of analytical tools
21
22 through which their concentration/activity can be speedily and easily measured at the time of
23
24 consultation^[8].
25
26
27
28
29
30

31 Despite an extensive candidate list, the diagnostic community has tended to focus on a
32
33 more limited group and an overview of the chemical components has been presented by
34
35 Harding et al^[8]. It must be noted however that although significant strides have been made in
36
37 terms of characterising the biochemical fluxes that can occur within a wound chronic wound
38
39 environment there is, as yet, no set of definitive markers.
40
41
42
43
44
45

46 5.0 Electrochemical Solutions

47 While there are a large number of systems that could ultimately be applied to wound
48
49 monitoring^[15], the present report has focused predominantly on electrochemical approaches.
50
51 The latter include: pH^[16-20], bacterial metabolites^[21-23], endogenous wound biomarkers^[24],
52
53 volatile organic emissions^[25,26] and temperature^[27]. Many of these have been investigated
54
55 with the intention of implementation within a clinic setting, but acquiring sufficient selectivity
56
57 with minimal sample preparation is a severe challenge, especially when considering the
58
59
60
61
62
63
64
65

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19
20
21
22
23
24
25
26
27
28
29
30
31
32
33
34
35
36
37
38
39
40

limited time available during a patient consultation. At present, considerable attention is being paid to the measurement of wound pH^[15-20]. The pH of the exudate within a wound site is known to vary according to the stage of healing and an idealized comparison of the pH profiles of acute and chronic wounds is highlighted in Figure 2. It is no surprise therefore that pH, will play a pivotal role in the biochemical reactions taking place and thus being able to monitor it could give the clinician valuable insights into the present state of the wound^[28]. While the natural pH of the skin will vary from person to person, it typically fluctuates within a narrow acidic range between pH 4 and pH 6 that impedes bacterial proliferation. Whenever an injury creates a cutaneous wound, it exposes the underlying tissue which, as it is normally regulated at pH 7.4, can encourage bacterial growth and thereby promote infection. In order to counter this, a temporary acidosis occurs during the initial healing stages^[28]. The pH of the chronic wound, however, oscillates in a weakly alkaline range and increases the susceptibility of the patient to bacterial incursions. Upon infection, bacteria seek to increase the wound pH to create an environment which is more accommodating to their growth^[28,29]. It has been proposed that this increase in pH could be a diagnostic handle through which to identify the onset of infection^[30] and there is an increasing effort to develop disposable pH sensors that could be applied to wound monitoring.

41
42
43
44
45
46
47
48
49
50
51
52
53
54
55
56
57
58
59
60
61
62
63
64
65

It is clear that the existing pH probe technology is far from suitable for use directly in assessing the pH of wound fluid but there have been considerable developments in recent years with a host of new approaches coming to the fore, largely in response to the limitations of the traditional glass potentiometric systems^[15,28,31]. It must be noted that not all are designed specifically for biomedical contexts, nor indeed are they all appropriate. The implementation of sensor systems within the latter gives rise to numerous concerns where issues over probe size and disposability can be problematic—especially where sample sizes may be limited or in vivo application is desired. A number of research avenues have been explored to counter these issues and both potentiometric^[31-44] and voltammetric^[16-20, 45-51]

1 methodologies have been pursued. In most cases, a functional material containing pH
2 sensitive components is applied to the surface of an appropriate sensor substrate through a
3 variety of methods that include: Adsorption/monolayer [39,48], polymer films [36,53], screen
4 printed inks [35,38,40,47], covalent attachment [46,47] or electrodeposition [32,33,37,42–44,50].
5
6
7
8

9
10 Screen printed electrochemical sensors have been designed specifically for wound pH
11 monitoring applications and shown to be a versatile addition^[16]. Rather than relying on
12 potentiometric detection methodologies, they have exploited voltammetric scanning (typically
13 using squarewave voltammetry) in which an endogenous biomarker such a uric acid (which is
14 ubiquitous within most fluids) is oxidised^[16-20]. The latter process is pH dependent and thus
15 the position of the oxidation peak can be used as indirect measure of pH. An example of the
16 signal output is highlighted in Figure 3. The single shot disposability proffered by standalone
17 screen printed systems does not however adequately meet the objectives of the Harding
18 Consensus document^[8] as it still requires the use of the device by an experienced healthcare
19 provider. It could be argued that although it could be exploited much in the same way as
20 glucose meters, in reality patient compliance will still be a major issue. A much better
21 approach would be to have a system that could periodically measure and autonomously report
22 back – either to the patient or to the clinician. This would require integrating the device within
23 the dressing itself as previously suggested. One innovative approach has been to examine the
24 use of woven carbon fibre as a sensing substrate^[17]. The carbon fibre weave is sufficiently
25 flexible to follow the contours of wounds that can present a highly variable surface
26 morphology and can be manufactured in the volumes necessary to enable the mass production
27 of dressing at relatively low cost. Carbon is already used in wound dressing and thus its
28 biocompatibility could, to a large extent, be assumed to have been proven. The main
29 challenge however relates to the activation of the fibres. The measurement methodology is
30 largely the same as for the screen printed systems^[16] in that urate is the target biomarker but
31 the basal plane nature of the fibres was found to be a considerable impediment to the
32
33
34
35
36
37
38
39
40
41
42
43
44
45
46
47
48
49
50
51
52
53
54
55
56
57
58
59
60
61
62
63
64
65

1 acquisition of a measurable voltammetric signal ^[17]. Atmospheric pressure plasma (APP)
2 treatment of the weave however has been investigated as a means of exfoliating the fibres
3 with the aim of improving the electron transfer kinetics – a prerequisite for analytical
4 applicability – and shown to significantly improve performance with a near nernstian
5 response obtained^[17].
6
7
8
9
10

11 The brittle nature of individual carbon fibres however present some additional issues
12 in that although flexible, fracturing and snapping can arise and fragmentation particles could
13 contaminate the wound and induce irritation leading to a prolonging of inflammation within
14 the wound site. The carbon fragments could also be embedded within the wound as tissue
15 remodeling commences leading to a need to debride the wound and further impeding the
16 healing processes. An alternative approach to counter this issue has been investigated through
17 the use of carbon composite polymers based on polyethylene or polycarbonate doped with
18 carbon particles^[18,19]. These films retain the flexibility of the weave without the fear of
19 fracture. More importantly, they can be directly integrated within a conventional wound
20 dressing. It was envisaged that the carbon component would provide the framework for
21 electrochemical transduction enabling quantitative information on key biomarkers associated
22 with wound healing to be extracted through an appropriate electronic monitor^[18] – worn either
23 by the patient or connected at the time of consultation. This would be a more viable solution
24 given the inherent mechanical flexibility of the composite film – a prerequisite given the high
25 degree of morphological variability encountered with diabetic ulcers. Moreover, the film
26 approach would have the critical advantage of component simplicity (requiring only carbon
27 and polyethylene) and hence provide a more inexpensive option when considering
28 manufacture. The possibility of extruding the material in the form of large area films would
29 be particularly amenable to large volume production – especially given the frequency with
30 which dressings would be changed. The electroanalytical performance of these films has
31
32
33
34
35
36
37
38
39
40
41
42
43
44
45
46
47
48
49
50
51
52
53
54
55
56
57
58
59
60
61
62
63
64
65

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19
20
21
22
23
24
25
26
27
28
29
30
31
32
33
34
35
36
37
38
39
40
41
42
43
44
45
46
47
48
49
50
51
52
53
54
55
56
57
58
59
60
61
62
63
64
65
been shown to accurately measure pH, having been tested in simulated and whole blood samples^[18,19].

Thus far, urate has been used as an indicator of pH yet it has been established in recent years that its quantification is of considerable interest as Fernandez et al.(2012) have shown that its concentration is elevated as a consequence of tissue remodeling within the wound. The cellular trauma and associated healing processes result in the release of adenosine triphosphate (ATP) into the extracellular wound matrix which is subsequently metabolised into uric acid^[24]. Thus, the ability to electrochemically monitor its concentration could be useful in determining the extent of inflammation and the subsequent response to treatment - therein providing insights into the healing process^[24]. This can be achieved through a procedurally simpler approach, involving the amperometric oxidation measurement of uric acid, As a biomarker, it has the advantage of being found in relatively high concentrations and there is an extensive amount of literature on its electrochemical detection^[27]. It is somewhat ironic to note that whilst the ease with which it can be electrochemically oxidised is often considered to be an interference in electroanalytical methods, it is this same property that has made it attractive as a biomarker in the present context. This approach has seen further refinement by Wang and co-workers who have adapted the system for wireless recording of the wound urate concentration through the use of an enzyme based system^[52].

6.0 Summary

Disposable screen printed electrodes have long been suggested as the preferred platform for point of care electrochemical sensors suitable for use within the clinic but while transfer to the home, in a manner analogous to personal glucose monitoring, would be ideal, the implementation is problematic. Chronic wounds invariably require handling under aseptic conditions by trained personnel and it is likely that having such processes conducted by a patient may interfere with the healing process, not only through mechanical disruption of the

1 wound but also by rendering the exposed tissue susceptible to infection. The “Connected
2 Health” scenario, where electrochemical techniques have the potential to excel, could be more
3 profitably exploited where the sensor substrates are directly integrated within the dressing.
4 These could then perform periodic monitoring of the wound status, typically pH, with
5 minimal intervention and report back to patient or, where appropriate, to a healthcare
6 professional in order effect a more speedy intervention. In many cases, simple amperometric
7 detection strategies could be employed but, even with more complex square wave techniques,
8 it is unlikely that the demands of the instrumentation will be an impediment to the
9 implementation of the technology. The challenge for realistic clinical success however will be
10 to provide economically viable, robust and clinically informative electrode systems that are
11 inherently disposable. This is particularly significant where sensors are embedded within
12 conventional dressings and may need to be changed daily over many months. It could be
13 anticipated that should these hurdles be overcome, the cost of the instrumentation would be
14 more than offset through avoiding the prolonged treatment and limb threatening
15 complications of infection.
16
17
18
19
20
21
22
23
24
25
26
27
28
29
30
31
32
33
34

35 Acknowledgements

36
37
38
39
40
41
42
43
44
45 The authors acknowledge the Department for Education and Learning Northern Ireland and
46 EC-Lab Ltd for funding this research.
47

48 Received: ((will be filled in by the editorial staff))

49 Revised: ((will be filled in by the editorial staff))

50 Published online: ((will be filled in by the editorial staff))
51
52
53

54 References

55
56
57 [1] J. Posnett, P. Franks, *Nursing Times*, **2008**, *104*, 44
58

59 [2] A.R. Siddiqui, J.M. Bernstein, *Clin Dermatol.* **2010**, *28*, 519.
60
61

- 1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19
20
21
22
23
24
25
26
27
28
29
30
31
32
33
34
35
36
37
38
39
40
41
42
43
44
45
46
47
48
49
50
51
52
53
54
55
56
57
58
59
60
61
62
63
64
65
- [3] N. Holman, R.J. Young, W.J. Jeffcoate, *Diabetologia*. **2012**, *55*, 1919.
- [4] L.I.F. Moura, A. Dias, E. Carvalho, H.C. Sousa, *Acta Biomaterialia*. **2013**, *7*, 7093.
- [5] V. Falanga, *The Lancet*. **2005**, *366*, 1736.
- [6] C. K. Sen, G. M. Gordillo, S. Roy, R. Kirsner, L. Lambert, T. K. Hunt, F. Gottrup, G. C. Gurtner, M. T. Longaker, *Wound Repair Regen*. **2009**, *17*, 763
- [7] K. Senarath-Yapa, S. Enoch *Wounds UK*, **2009**, *5*, 38
- [8] K. Harding, A World Union of Wound Healing Societies, “Principles of best practice: Diagnostics and wounds: A consensus document”, London: MEP Ltd, 2008.
- [9] S. Schreml, R.M.Szeimies, L. Prantl, M. Landthaler, P. Babilas, *J. Am. Acad. Derm.* **2010**, *63*, 866.
- [10] N.C. Schaper, J. Apelqvist, K. Bakker, *Curr. Diab. Rep.*, **2003**, *3*, 475.
- [11] M.W. Löffler, H. Schuster, S. Bühler, and S. Beckert *The International Journal of Lower Extremity Wounds*, **2013**, *12*, 113
- [12] M. Schmohl, S. Beckert, T.O. Joos, A. Königsrainer, *Diabetes Care*, **2012**, *35*, 2113
- [13] R. Zillmer, H. Trøstrup, T.Karlsmark, P. Ifversen, M.S.Agren *Arch Dermatol Res*.**2011**, *303*, 601
- [14] J.Broadbent, T. Walsh, Z. Upton. *Proteomics Clin. Appl.* **2010**, *4*, 204
- [15] T.R. Dargaville, B.L.Farrugia, J.A.Broadbent, S. Pace, Z. Upton, *Biosens. Bioelectron.* **2013**, *41*, 30.
- [16] J. Phair, L. Newton,C. McCormac, M.F. Cardosi, R. Leslie, J. Davis, *Analyst*. **2011**, *136*, 4692.
- [17] J. Phair, C.P.Leach, M.F.Cardosi, J. Davis, *Electrochem. Commun.* **2013**, *33*, 99.
- [18] J. Phair, M. Joshi, J. Benson, D. McDonald, J. Davis, *Materials Chemistry and Physics*, **2014**, *143*, 991
- [19] J. Phair, J. Benson, C. McCormac, J. Cundell, S. Gracheva, D. Wilkinson, S. Forsythe, J. Davis, *Sensors and Actuators B*, **2014**, *193*, 764-7

- 1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19
20
21
22
23
24
25
26
27
28
29
30
31
32
33
34
35
36
37
38
39
40
41
42
43
44
45
46
47
48
49
50
51
52
53
54
55
56
57
58
59
60
61
62
63
64
65
- [20] A. McLister, J. Davis, *Healthcare*, **2015**, 3, 466.
- [21] D. Sharp, P. Gladstone, R. Smith, S. Forsythe, J. Davis, *Bioelectrochemistry*. **2010**, 77, 114.
- [22] L. Zhou, J.D. Glennon, J.H.T. Luong, F.J. Reen, F. O'Gara, C. McSweeney, G.P. McGlacken, *Chem. Commun.* **2011**, 47, 10347.
- [23] I. Ciani, H. Schulze, D.K. Corrigan, G. Henihan, G. Giraud, J.G. Terry, A.J. Walton, R. Pethig, P. Ghazal, J. Crain, C.J. Campbell, T.T. Bachmann, A.R. Mount, **Biosens. Bioelectron.** **2012**, 31, 413.
- [24] M.L. Fernandez, Z. Upton, H. Edwards, K. Finlayson, G.K. Shooter, *International Wound Journal*. **2012**, 9, 139.
- [25] J. Feng, F. Tian, J. Yan, Q. He, Y. Shen, L. Pan, *Sens. Actuators B.* **2011**, 157, 395.
- [26] A.L.P.S. Bailey, A.M. Pisanelli, K.C. Persaud, *Sens. Actuators B.* **2008**, 131, 5.
- [27] G. Matzeu, A. Pucci, S. Savi, M. Romanelli, F. Di Francesco, *Sens. Actuators A.* **2012**, 178, 94.
- [28] L.A. Schneider, A. Korber, S. Grabbe, J. Dissemond, *Arch. Derm. Res.* **2007**, 298, 413.
- [29] V. Sridhar, K. Takahata, *Sens. Actuators B.* **2009**, 155, 58.
- [30] L. Shi, S. Ramsay, R. Ermis, D. Carson, *J. Wound Ostom. Cont.* **2011**, 38, 514.
- [31] G.J.M. Hemmink, B.L.A.M. Weusten, J. Oors, A.J. Bredenoord, R. Timmer, A.J.P.M. Smout, *Eur. J. Gastroenterol. Hepatol.* **2010**, 22, 572
- [32] A. Taouil, F. Lallemand, J.M. Meolt, J. Husson, J.Y. Hihn, B. Lakard, *Synth. Met.* **2010**, 160, 1073
- [33] S. Carrol, R.P. Baldwin, *Anal. Chem.* **2010**, 82, 878
- [34] G. Herlem, R. Zeggari, J.Y. Rauch, S. Monney, F.T. Anzola, Y. Guillaume, C. Andre, T. Gharbhi, *Talanta* **2010**, 82, 417

- 1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19
20
21
22
23
24
25
26
27
28
29
30
31
32
33
34
35
36
37
38
39
40
41
42
43
44
45
46
47
48
49
50
51
52
53
54
55
56
57
58
59
60
61
62
63
64
65
- [35] A.E. Musa, F.J. del Campo, N. Abramova, M.A. Alonso-Lomillo, O. Dominguez-Renedo, M.J. Arcos-Martinez, M. Brivio, D. Snakenborg, O.Geschke, J.P. Kutter, *Electroanalysis* **2011**, *23*, 115
- [36] T. Inoue, T. Baba, A. Yuchi, A. *Electroanalysis* **2011**, *23*, 536
- [37] Q. Li, H. Li, J. Zhang, Z. Xu, *Sens. Actuators B* **2011**, *155*, 730
- [38] S. Betelu, K. Polychronopoulou, C. Rebholz, I. Ignatiadis, I. *Talanta* **2011**, *87*, 126
- [39] D. Lee, T. Cui, *Microelectron. Eng.* **2012**, *93*, 39
- [40] S. Zhuiykov, E. Kats, K. Kalantar-Zadeh, M. Breedon, N. Miura, *Mater. Lett.* **2012**, *75*, 165
- [41] S. Kim, T. Rim, K. Kim, U. Lee, E. Baek, H. Lee, C.K. Baek, M. Meyyappan, M.J. Deen, J.S. Lee, *Analyst* **2011**, *136*, 5012
- [42] N. Cherchour, C. Deslouis, B. Messaoudi, *Electrochim. Acta* **2011**, *56*, 9746
- [43] R. Zhao, M. Xu, J. Wang, G. Chen, *Electrochim. Acta* **2010**, *55*, 5647
- [44] A. Mignani, C. Corticelli, D. Tonelli, E. Scavetta, *Electroanalysis* **2011**, *23*, 1745
- [45] J.J. Hickman, D. Ofer, P.E. Laibinis, G.M. Whitesides, M.S. Wrighton, *Science* **1991**, *252*, 688
- [46] A. Makos, D.M. Omiatek, A.G. Ewing, M.L. Heien, *Langmuir* **2010**, *26*, 10386
- [47] L. Xiong, C. Batchelor-McAuley, R.G. Compton, *Sens. Actuators B* **2011**, *159*, 251
- [48] W. Park, S. Kim, *Electrochem. Commun.* **2013**, *26*, 109
- [49] I. Streeter, H.C. Leventis, G.G. Wildgoose, M.Pandurangappa, N.S. Lawrence, L. Jiang, T.G. Jones, R.G. Compton, *J. Solid State Electrochem.* **2004**, *8*, 718
- [50] J. Davis, M.T. Molina, C.P. Leach, M.F. Cardosi, *ACSApp. Mat. Interfaces* **2013**, *8*, 9367
- [51] M. Li, J. Phair, M.F. Cardosi, J. Davis, *Anal. Chim. Acta* **2014**, *812*, 1
- [52] P. Kassala, J. Kima, R. Kumara, W.R. de Araujo, I. Murković Steinberg, M.D. Steinberg, J. Wang, *Electrochem. Commun.*, **2015**, *56*, 6




Figure 1. Interdisciplinary nature of smart dressing research

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19
20
21
22
23
24
25
26
27
28
29
30
31
32
33
34
35
36
37
38
39
40
41
42
43
44
45
46
47
48
49
50
51
52
53
54
55
56
57
58
59
60
61
62
63
64
65




Figure 2. Typical pH profiles observed for acute and chronic wounds.
(Adapted from Scheinder *et al.*2007[28])





Figure 3. Different material approaches to the design of conductive smart dressings: carbon fibre weave (A), laser etched carbon loaded polyethylene mesh(B), and carbon-polycarbonate film integrated within a conventional dressing (C). The potential implementation of the dressing is indicated in the sketch in (D) and the typical voltammetric pH profile for the oxidation of uric acid is highlighted in (E) [16-20]


1
2
3
4
5
6
7
8
9
10
11 **The availability of sensors that could measure wound pH, either in the clinic or in home**
12 **could significantly improve clinical outcome – particularly in the early identification of**
13 **complications such as infection.** This review identifies new materials and electrochemical
14 research strategies that are being targeted at wound diagnostics and the design of smart
15 dressings
16

17
18 **Keywords: Electrode;sensors;Wound;Healing;Infection**
19



43
44 Anna McLister¹, Jolene Phair¹, Jill Cundell², James Davis^{1*}
45

46 **New Developments in Smart Bandage Technologies for Wound Diagnostics**
47
48
49
50
51
52
53
54
55
56
57
58
59
60
61
62
63
64
65




Click here to access/download

Production Data

Figure 1.jpg






Click here to access/download

Production Data

Figure 2.jpg






Click here to access/download

Production Data

Figure 3.jpg





Click here to access/download

Production Data
Graphic Abstract.jpg

