



LSTM Classification of sEMG Signals For Individual Finger Movements Using Low Cost Wearable Sensor

Millar, C., Siddique, N., & Kerr, E. (2020). LSTM Classification of sEMG Signals For Individual Finger Movements Using Low Cost Wearable Sensor. In *2020 International Symposium on Community-Centric Systems, CcS 2020* (pp. 1-8). Article 9231515 (2020 International Symposium on Community-Centric Systems, CcS 2020). IEEE. <https://doi.org/10.1109/CcS49175.2020.9231515>

[Link to publication record in Ulster University Research Portal](#)

Published in:

2020 International Symposium on Community-Centric Systems, CcS 2020

Publication Status:

Published (in print/issue): 20/10/2020

DOI:

[10.1109/CcS49175.2020.9231515](https://doi.org/10.1109/CcS49175.2020.9231515)

Document Version

Author Accepted version

General rights

The copyright and moral rights to the output are retained by the output author(s), unless otherwise stated by the document licence.

Unless otherwise stated, users are permitted to download a copy of the output for personal study or non-commercial research and are permitted to freely distribute the URL of the output. They are not permitted to alter, reproduce, distribute or make any commercial use of the output without obtaining the permission of the author(s).

If the document is licenced under Creative Commons, the rights of users of the documents can be found at <https://creativecommons.org/share-your-work/licenses/>.

Take down policy

The Research Portal is Ulster University's institutional repository that provides access to Ulster's research outputs. Every effort has been made to ensure that content in the Research Portal does not infringe any person's rights, or applicable UK laws. If you discover content in the Research Portal that you believe breaches copyright or violates any law, please contact pure-support@ulster.ac.uk

LSTM Classification of sEMG Signals For Individual Finger Movements Using Low Cost Wearable Sensor

Christopher Millar

Faculty of Computing, Engineering and Built Environment

Ulster University

Derry, N. Ireland

millar-C21@ulster.ac.uk

Dr. Nazmul Siddique

Faculty of Computing, Engineering and Built Environment

Ulster University

Derry, N. Ireland

nh.siddique@ulster.ac.uk

Dr Emmett Kerr

Faculty of Computing, Engineering and Built Environment

Ulster University

Derry, N. Ireland

ep.kerr@ulster.ac.uk

The electrical activity of the muscles that control finger movements can be extracted during the performance of these movements and using machine learning techniques, the myoelectric signals can be decoded and classified according to the movement that generated the specific signal. The focus of this paper is to classify sEMG signal using easily accessible cheap hardware to capture the signal. Furthermore, to employ neural networks to classify the signal using established methodology i.e. feature extraction, with the highest possible accuracy. To classify these sEMG signals, an LSTM network has been developed and was able to classify 12 individual finger movements with accuracies reaching 90%.

Keywords—sEMG, LSTM, Myo, Finger Movement Classification

I. INTRODUCTION

Endowing robots with grasping capabilities similar to the dexterity demonstrated by humans is a massive challenge [1], yet it will provide great autonomy to robots and will allow them to interact with humans in a natural way. Robotic grasping is a highly complex problem which requires knowledge about the shape, mass distribution, and friction of the object to grasp, but is also conditioned by the purpose of the proposed task to be executed with the object. Standard approaches to robot grasping try to take all these aspects into consideration and devise a plan on how to move individual fingers to achieve a suitable grasp. A promising alternative consists on learning grasping movements and configurations from human examples. However, in order to train a robotic grasping system using human demonstrations is necessary to know precisely how a human performs this task. This paper presents the contribution towards human demonstrations for robotic grasping by identifying different finger movement using a low-cost electromyography system.

A popular method of examining human body movements in biomechanics is through the electrical outputs of the muscles involved in the movements being investigated, known as electromyography (EMG) [2]–[5]. One technique for measuring the EMG signal is using intramuscular needles inserted into the muscles to detect the myoelectric signal. Although very accurate, this method is a highly invasive

procedure that requires a medical expert to ensure the safety of the subjects and correct placement of the electrodes on the muscles [6]. Another approach is through surface electromyography (sEMG), a non-invasive method that involves placing an electrode on the surface of the skin above the muscle regions to detect the muscles electrical activity. A comparative study concluded that there is no significant difference in the accuracy of the classification, therefore sEMG is generally preferred due to its non-invasive nature [6].

The sEMG signal has been used in models that have been developed for classifying individual finger movements with a high degree of accuracy [4], [7]–[9]. Through the development of feature extraction techniques in combination with machine learning methods, improved performance of sEMG signal classification has been achieved which makes it a suitable technique for application with robotic systems [10]. Applied to robotics it can be an intuitive method of controlling or training an anthropomorphic robotic system based on real world human demonstrations.

II. RELATED WORK

In [2], [11] and [12] a small selection of basic movements were performed i.e. simple directional arm movements up, down, left & right [9], elbow flexion and extension and forearm rotation [2] as well as functional gestures: power grip, precision grip, open hand, pointed index finger and wrist flexion/extension. The classification methods varied between these papers with [2] and [9] employing a back propagation neural network (BPNN), whereas [14] used an SVM classifier. A selection of commonly found time and frequency domain features were extracted from the original EMG signal in all these papers. All these works investigate similar type of movements, movements that generate larger muscle contractions, and therefore generate signals with larger signal to noise ratio. Furthermore, these papers use high sampling ratio e.g. 1000Hz devices, for collecting the data. This has been shown to be beneficial when performing sEMG signal classification as they can generate more data per movement making it easier to discriminate between the movement classes [13], [14].

Modelling finger movements through decoding EMG signals generated by the muscles of the forearm when contracting to perform flexion and extension of the digits is a

relatively new application of existing EMG based techniques. In [4] the authors used an array of 32 sEMG electrodes wrapped around the forearm of the user to collect the sEMG data of a total of 10 finger movements i.e. flexion/extension of each finger and 2 grouped movements of the middle, ring and little fingers. The detection device sampled data at a frequency of 2000Hz and four commonly found time domain features were extracted from the original signal and used as inputs into a feedforward neural network (FNN). To improve the classification of the network the features were extracted from overlapping windows. In a follow up work the authors classified individual finger movements of amputee subjects using the same protocols [7]. The research presented in [9] revolved around a benchmark database: Ninapro¹ where the authors collected data from able bodied and amputee subjects using different sEMG technologies whilst performing a benchmark set of movements that are associated with activities of daily living [15]. The authors collected sEMG signals using a medical grade device with ten electrodes and a sample frequency of 100Hz and positioned the electrodes uniformly around the subjects forearm. A Support Vector Machine (SVM) was utilized to classify 12 different movements, ten of which were individual finger movements, i.e. extension/flexion index finger. Although the author does not appear to employ any form of signal segmentation which could account for not achieving higher classification of the signals but as they are using a detection device with a large sampling rate and investigating movements that generate large signals they are still able to produce results with high levels of accuracy [9]. Using this medical grade equipment the authors were able to successfully classify with a relatively high degree of accuracy, the authors also noted that this is a commercial device that records a rectified signal rather than raw data and that this can have an effect on the final classification accuracy. These studies yielded classification accuracies between 80-90% for individual finger movements, however, they employed expensive medical grade EMG detection systems [9] or detection systems with high sample rate [4]. Systems like these are not always a viable option for researchers due to the cost of the systems and can require a level of expertise to correctly setup and operate.

The introduction of commercially available cheap wearable sensors offers an alternative method of sEMG signal capturing. Recently the Myo gesture control armband² has been released which has multiple applications i.e. a controller for an intuitive music device [16], sign language translation [17] and controlling robots and prosthetic limbs [2]–[4], [9], [11], [18]. However, these devices typically have a lower sampling rate (e.g. Myo gesture control armband samples at 200Hz) and a lower signal to noise ratio making the signal classification problem more difficult. The development of an intuitive system in conjunction with easily accessible wearable sensors could lead to a sophisticated grasping model that is based on real world human demonstration. Such a system will then dictate the movements performed by an anthropomorphic robotic hand

when picking up objects and performing tasks associated with activities of daily living. By measuring and collecting the data of the electrical impulses with one of these commercial devices and combining it with state-of-the-art machine learning algorithms e.g. neural networks, and robust feature selection, the sampling rate drawback can be mitigated against and a high level of accuracy can still be achieved [16], [19]–[22].

Research using the Myo armband has shown how neural networks have been applied to sEMG signal classification i.e. convolutional neural networks (CNN) and Long Short Term Memory (LSTM) networks [8], [19], [23]. Stephenson et al (2018), used the sEMG signal pattern as an image input into a CNN and performed classification on five finger flexion movements as well as seven gestures which included four combinations of flexion movements involving the thumb and one of the other fingers i.e. index [8]. In [23], LSTM and CNN networks were employed as the method of signal classification and were compared against each other. A bespoke architecture that combined LSTM layers with a single CNN layer that was named LSTM-CNN (LCNN). The authors recorded their own dataset that included some basic hand and wrist gestures i.e. wrist extension & closed fist. In [19], using an LSTM network and a mixture of different domain features i.e. Time & Frequency, they were able to classify some basic hand gestures i.e. open hand and closed fist. Further research has been carried out in [17], [24], [25] using the Myo armband. [17], used a CNN to classify signals extracted from the subjects when performing any of 30 sign language gestures. In [24] six basic hand gestures e.g. open hand and cylindrical grasp, were classified using a set of five time domain features extracted from overlapping windows and input into an SVM classifier. The gestures involved in these research studies require coupled movements of the fingers and arms which can generate signals with greater amplitudes than that of the signals generated by individual finger movements.

As demonstrated in the majority of the papers that have been reviewed, classification of hand gestures can be achieved using a wide range of classification methods e.g. SVM, LDA, ANN, LSTM, and CNN. This can be down to the fact that the generated sEMG signal of these dynamic gestures i.e. wrist flexion or gestures involving multiple fingers, generally generate signals with more pronounced amplitudes. Whereas, with individual finger movements the detected signal will be a lot more difficult to distinguish between movement classes due to the reduced amplitudes of the already inherently weak signals that are generated and cross over where the same muscle control multiple fingers i.e. mid, ring and little fingers.

Furthermore, a range of different devices have been evaluated in the reviewed studies and it is evident that sEMG signal capture systems with higher sampling rates have demonstrated the ability to discriminate subtle individual finger movements. Whereas, devices with lower sampling rates, e.g. Myo armband, are less able to complete the same task. Another factor that has been highlighted is that when more movement

¹ <http://ninaweb.hevs.ch/>

² <https://support.getmyo.com/hc/en-us/articles/203398347-Getting-started-with-your-Myo-armband>

classes are introduced for classification the final classification accuracy is often reduced. This affects the classification of all devices but more so devices with reduced sampling rates.

The focus of this paper is to develop a neural network that can classify individual finger movements from the sEMG signal detected using wearable EMG detection hardware. Figure 1 shows an abstract view of the pipeline of the proposed system that is to be developed as part of the research being conducted.




Figure 2: Abstract view of proposed system architecture

The rest of this paper is organised as follows: Section III outlines the technology being used along with the architecture of the neural network that was created, section IV details the experimental protocols that were followed. Section V details the results of the experiments and finally a conclusion in section VI.

III. METHODOLOGY

Figure 1 illustrates an abstract overview of the proposed pipeline of the system and this following section will describe each stage of the pipeline. Section A provides a detailed description of the technology being used for data collection: the Myo armband. Section B focuses on feature extraction techniques. The final section will describe the LSTM network architecture used for the classification of the movement signals.

A. sEMG Device

The Myo armband, shown in Figure 2(a), is a wireless device that detects the inherently weak EMG signal generated by the forearm muscles in addition to other spatial information e.g. Orientation. The Myo is made up of eight medical grade stainless steel dry electrodes that detect the electrical output of the muscles found in the forearm that are responsible for the dexterous movement of the fingers and thumb i.e. Flexor Digitorum Profundus or Extensor Digitorum Communis. The Myo armband represents the EMG signal and normalises it to within a value range between -128 and 128. These values are the amplified detected signal that is measured by the eight electrodes. The Myo armband is also equipped with a nine axis inertial measurement unit (IMU) that contains a three axis gyroscope, three axis accelerometer and a three-axis magnetometer that can measure the speed of movements and orientation of the armband in 3D space. The addition of these extra sensors allows for the potential fusion of sEMG and IMU data from a wearable device. By fusing the data provided by the multiple sensors within the Myo it could allow for more advanced autonomous control systems by giving the system more information to analyse and learn from [13], [14], [19]. The main drawback with using the Myo is the fact that it has a sample rate of 200Hz which is lower than medical grade sensors that are used in other sEMG signal classification papers e.g. [7], [18], [20], [21]. However, whilst this is a valid concern the

difference between the overall classification accuracy when using the devices with a lower sampling frequency i.e. Myo, has been demonstrated to be less than 5% when employing techniques like feature extraction and sliding windows to reduce the impact of the low sampling rate [13], [14].



Figure 1: (a) Myo Armband (b) Subject wearing device around forearm

B. Feature Extraction

A large number of possible features used to reduce the dimensionality of the dataset can be found in the literature. These features can be split into three common domains: Time, Frequency and Time-Frequency. Features within these domains have all been used as a method of improving sEMG signal classification accuracy. Time domain features are the most commonly used throughout the literature as they are the most efficient in terms of calculation time and classification result. The initial set of time domain features used are the most commonly used in sEMG signal classification, these features will be tested individually and within different sets to find the features that produce the best classification performance. The features extracted from the EMG signals in this research are:

- *Mean Absolute Value (MAV)* is the most commonly used feature found in the literature. It is the average of the absolute value of the EMG signal for each window that the signal has been segmented into. X is the MAV of the signal in segment i which is N samples long. X_k is the k^{th} sample in segment i and I is the total number of segments that the original signal sample has been split into.

$$\bar{x}_i = \frac{1}{N} \sum_{k=1}^N |x_k| \text{ for } i = 1, \dots, I \quad (1)$$

- *Waveform Length (WL)* is another popular feature from the time domain that has been used throughout the literature. This feature is used to quantify the complexity of the waveform in each signal segment. It is the cumulative length of the waveform over the entire signal segment. $X_{n+1} - X_n$ is the difference in consecutive sample voltage values.

$$WL = \sum_{n=1}^{N-1} |x_{n+1} - x_n| \quad (2)$$

- *Variance (VAR)* has been used as a time domain feature in gesture recognition studies. Variance is a measure of the

power of an EMG signal and represents the deviation of the EMG signal from its mean value.

$$\text{VAR} = \frac{1}{N-1} \sum_{n=1}^N x_n^2 \quad (3)$$

- *Slope Sign Change (SSC)* has been extensively used in the literature, it represents the number of times that the EMG signal changes between the positive and negative slope changes between three sequential segments. The threshold is normally selected between 50 μV and 100 mV.

$$\text{SSC} = \sum_{n=2}^{N-1} [f[(x_n - x_{n-1}) \times (x_n - x_{n+1})]] \quad (4)$$

- *Autoregressive Modelling (AR)* is used for classification of gestures and finger movements as it has been shown that the sEMG spectrum changes when muscle contractions are performed. a_i are the AR coefficients, p is the model order and e_k is the residual white noise. Different coefficients have been found to be suitable for feature extraction of EMG signal ranging from 1 to 10 [13], [26]–[29]. In this paper the 1st and 2nd orders are used.

$$X_k = \sum_{i=1}^p a_i X_{k-i} + e_k \quad (5)$$

- *Zero Crossings (ZC)* represents frequency information of an EMG signal but it is defined in the time domain. It is the number of times that the amplitude of the EMG signal crosses the zero amplitude level. As a way of avoiding signal fluctuations or signal noise a threshold condition is implemented.

$$\text{ZC} = \sum_{n=1}^{N-1} [\text{sgn}(x_n, x_{n+1}) \cap |x_n - x_{n+1}| \geq \text{threshold}] \quad (6)$$

$$\text{Sgn}(x) = \begin{cases} 1 & \text{if } x \geq \text{threshold} \\ 0 & \text{otherwise} \end{cases}$$

- *Willison Amplitude (WAMP)* is the resulting number of times that the difference between two adjacent signal segments amplitude exceeds a pre-set threshold as a method of reducing noise similar to ZC and SSC.

$$\text{WAMP} = \sum_{n=1}^{N-1} f(|X_n - x_{n+1}|) \quad (7)$$

$$f(x) = \begin{cases} 1 & \text{if } x \geq \text{threshold} \\ 0 & \text{otherwise} \end{cases}$$

- *Root Mean Square (RMS)* this is the square root of a signal segment's mean value. This feature has been used in combination with other time domain features in many sEMG signal classification models.

$$\text{RMS} = \sqrt{\frac{1}{N} \sum_{n=1}^N x_n^2} \quad (8)$$

- *Standard Deviation (STD)* measures the total variation of a particular set of values against the mean of the population.

This time domain feature has been used successfully in combination with other time domain features[11].

$$\sigma = \sqrt{\frac{1}{N} \sum_{i=1}^N (x_i - \mu)^2} \quad (9)$$

- *Mean Absolute Deviation (MAD)* is a measure of the average distance between the mean of a dataset and each data value within it. This feature was used by [30] when using the Myo as part of the authors hand pose recognition system.

$$\text{MAD} = \frac{1}{N} \sum_{i=1}^n |x_i - m(X)| \quad (10)$$

- *Kurtosis (KURT)* is used as measure of how outlier-prone a distribution of a dataset is.

$$\text{Kurt} = \frac{\frac{1}{n} \sum_{i=1}^n (x_i - \bar{x})^4}{(\frac{1}{n} \sum_{i=1}^n (x_i - \bar{x})^2)^2} \quad (11)$$

A technique that can be employed to enhance the effectiveness of feature extraction is the use of overlapping sliding windows [4], [7], [9], [31], [32]. Overlapping windows are used as a method of retaining as much information as possible when extracting features but also for reducing the dimensionality of the original signal. By using overlapping windows the amount of data that is extracted from the signal is increased which potentially increases the classification accuracy of the proposed system. Selection of the window size has been chosen based on previous works found in the literature, the most common window size found was 200ms [4], [7], [9], [14], [32]. A window overlap of 20ms, has been selected to mitigate the reduced sample rate of the Myo device and provide as much data from the reduced original signal as possible to the LSTM classifier.

C. Classification using LSTM Network

LSTM is a recurrent neural network (RNN) that is trained through a gradient based learning algorithm that was introduced as a solution to the problems with error block-flow found in other “Back Propagation Through Time” (BPTT) and “Real-Time Recurrent Learning (RTRL)” networks. The LSTM network forces constant error flow through the internal states of the special units found within the LSTM network architecture [33]. The introduction of multiplicative input gate units and output gate units allows for constant error flow. These gated cells control the flow of data depending on its strength and importance by either passing the contents on to the next cell or by blocking the information. This mechanism is controlled by the modification of the weights through the networks learning process [33]. Figure 3, shows an overview of the architecture that was used to classify sEMG signals in this work.

The LSTM network that has been developed for this work is comprised of six different layers. The initial input layer is a sequence input layer that allows sequential data to be input into the network. Each movement sequence sample is made up of an input vector that represents each feature that has been extracted from the original signal of each of the eight individual sEMG sensors. To further optimize the LSTM network the input data was normalized using the Z-score of the input [34]. This adheres to the protocol followed in [19], [23].

The sequences are input into a bidirectional LSTM (bi-LSTM) layer which enables the network to learn bi-directional long-term dependencies between the different time steps of the sequential data allowing the network to predict using the entire sequential input [22]. The number of hidden units in the bi-LSTM layer was empirically tested during the experiment to find the configuration that produced the highest classification accuracy.

A dropout layer was added after the bi-LSTM to reduce overfitting [23]. Dropout is a method of network regularization that attempts to reduce the co-adaptation of the hidden units within the network. The dropout layer operates by randomly deactivating hidden units with a probability of p , in this case $p = 0.3$ [23]. The fourth layer in the network is the fully connected layer that multiplies the inputs by a weight matrix and adds a bias vector to the input, this layer then feeds into the softmax layer where a softmax function is applied to the activation. Finally there is a classification layer that computes the cross entropy loss for classification networks with mutually exclusive classes. Further parameters can be adjusted to suit the particular problem that is being investigated e.g. mini batch size, number of hidden units, additional layers.




Figure 3: LSTM Architecture

IV. DATA ACQUISITION AND PRE-PROCESSING

This section describes the protocols followed during the experimentation process will be outlined. The main focus will be on how the data was collected and the pre-processing methods that were applied to the EMG signal.

A. Data Acquisition Protocol

The Myo was placed securely around the widest part of the subjects forearm. The subjects left arm was placed on a small platform, 16cm in height, in an effort to remove any muscle activity that would be used to hold arm in the air, as shown in Figure 2 (b). The hand was in a relaxed position with no contact being made with the table below. Flexion and extension of the fingers and thumb along with abduction and adduction of the thumb were performed seventy times by the subject along with a rest pose. The samples were randomly split into two sets, one

for training containing 728 (80%) samples and a test set containing 182 (20%) samples of each movement [17]. A script was created in Matlab³ to prompt the subject to begin the movement and when to return to rest position. An initial 4 second rest period began each trial, once the user was prompted to perform the movement the pose was held for 5 seconds and then returned to rest position to complete the data collection process. Using Matlab and a specifically designed toolbox, Myo SDK MATLAB MEX wrapper [35], the raw signal data was extracted from the Myo armband and imported into Matlab as the subject was performing individual finger movements i.e. Index extension/flexion, as shown in Figure 4.



Figure 4: Finger Poses conducted for Classification. From (Top Row) Left to Right: Little Extension; Flexion; Ring Extension; Flexion; Mid Extension; Flexion; (Bottom Row) Index Extension/ Flexion; Thumb Extension; Flexion; Thumb Abduction; Adduction

B. Data Pre-Processing

The Myo armband transmits the detected signals over Bluetooth via the Bluetooth dongle to the PC. The Myo has a built in notch filter that filters out signals at 50Hz, this is done to remove any interference with the EMG signal caused by European power line interference. This interference is caused by electrical interference of other systems e.g. Myo's battery, with the EMG signal. No other filtering was completed on the recorded signal. The next stage of the protocol involved removing the surplus data. As aforementioned, the movement data started after the initial 4 second rest period and ended 5 seconds later. All data before 3.5 seconds was removed, this allowed a small 0.5 second buffer to allow for user error in case of the action starting before the 4 second mark. Cropping the samples to a fixed consistent length is also important when working with LSTM networks, as with other types of RNN's, as they require the input sequence samples to be of the same length. LSTM will automatically pad all samples that are shorter than the longest sequence in the data set. The addition of too much padding can distort the signals and therefore can lead to a reduction in classification accuracy [36]. Figure 5 shows an example of each EMG channel's signal when performing the different finger movements.

V. RESULTS

An initial set of experiments were conducted where a network was created using a single set of features that included all the time domain features discussed in Section III.B. The average result from 30 trained LSTM networks when classifying 13 movements (12 index/flexion movements & 1 rest pose) along with the best single performance of a single network was

³ <https://uk.mathworks.com/>

recorded. Table 1 shows the parameters used for the LSTM network that will remain the same throughout the testing.

Table 1: LSTM Base Parameters

EPOCH	Mini Batch	Normalization	Dropout	Weights	No of Movements	Features
25	56	ZSCORE	0.3	GLOROT	13	ALL

To improve the classification accuracy of the network being implemented in this current research the number of hidden units in each layer were incrementally adjusted and additional bi-LSTM layers were added to find the best combination that produces the highest classification accuracy possible. The number of hidden units was incrementally increased until the average classification accuracy no longer increased. In the figure below, Figure 5, it shows the average classification for each of the networks trained when incrementing the number of hidden units along with the result of the best performing network.




Figure 5: Results of LSTM with single bi-LSTM layer when incrementing number of hidden units

As shown in Figure 5 there is a steady increase in the average accuracy that peaks when using 180 hidden units in the bi-LSTM layer. The average accuracy achieved here is 84.49%, however, the single best performing network that uses 180 hidden units achieved an accuracy of 87.36%. There were other networks that achieved a higher single accuracy e.g. 87.91%, when using less hidden units, 160 or 170, but the differences were negligible and the highest average was used as the benchmark architecture.




Figure 6: Results of LSTM with two bi-LSTM layer when incrementing number of hidden units

A second bi-LSTM layer was added to the network with a dropout layer and the hidden units for that layer were

incremented again, starting at 10 hidden units, until there was no more improvement on the average classification accuracy. Figure 6, above, shows that when using 2 bi-LSTM layers with 180 and 60 hidden units in the first and second layers respectively, the average accuracy was increased to 86.76% with single best accuracy of 89.56% occurring with this LSTM architecture.

A third layer was added and the same procedure was followed, the number of hidden units was incremented until the average accuracy had peaked. Figure 7 shows that the performance of the networks, when a third layer is added, never improves upon the two layer setup. Whilst some of the individual networks achieve an accuracy of 89.56% the average accuracies all drop below 86% so test up to 50 hidden units were only carried out to confirm the continual drop in classification performance.




Figure 7: Results of LSTM with three bi-LSTM layer when incrementing number of hidden units

The overall best performing LSTM architecture from these empirical studies were with two bi-LSTM layers and with 180 hidden units in the first layer and 60 in the second. With this architecture the single best performing network was 89.56%. Nonetheless, there were other architectures that classified with higher accuracies, e.g. 180 hidden units with either 10 or 80 units in the additional bi-LSTM layers, the increase is insignificant at less than 1%.

		Accuracy: 89.56%													
Output Class	IND EXT	85.7%	0.0%	0.0%	0.0%	0.0%	0.0%	0.0%	0.0%	0.0%	5.6%	0.0%	0.0%	0.0%	0.0%
	IND FLX	19	94.4%	5.6%	0.0%	0.0%	0.0%	0.0%	0.0%	0.0%	0.0%	0.0%	0.0%	0.0%	0.0%
	MID EXT	0	0	88.9%	0.0%	0.0%	0.0%	0.0%	0.0%	0.0%	0.0%	0.0%	0.0%	0.0%	0.0%
	MID FLX	0	0	16	100.0%	0.0%	0.0%	0.0%	0.0%	0.0%	0.0%	0.0%	0.0%	0.0%	0.0%
	RING EXT	0	0	0	14	100.0%	0.0%	0.0%	0.0%	0.0%	0.0%	0.0%	0.0%	0.0%	0.0%
	RING FLX	0	0	1	0	16	100.0%	0.0%	0.0%	0.0%	0.0%	0.0%	0.0%	0.0%	0.0%
	LIT EXT	0	0	0	0	11	100.0%	0.0%	0.0%	0.0%	0.0%	0.0%	0.0%	0.0%	0.0%
	LIT FLX	0	0	0	0	10	100.0%	0.0%	0.0%	0.0%	0.0%	0.0%	0.0%	0.0%	0.0%
	THMB ADD	14.3%	0.0%	0.0%	0.0%	0.0%	0.0%	0.0%	0.0%	0.0%	94.4%	0.0%	0.0%	21.4%	0.0%
	THMB ABD	3	0	0	0	0	0	0	0	0	17	100.0%	0.0%	3	0
	THMB EXT	0	0	0	0	0	0	0	0	0	72.0%	88.9%	0.0%	0	0
	THMB FLX	0	0	0	0	0	0	0	0	0	3	5	61.5%	7.1%	0
	REST	0	0	0	0	0	0	0	0	0	0	0	0	10	100.0%
	IND EXT	IND FLX	MID EXT	MID FLX	RING EXT	RING FLX	LIT EXT	LIT FLX	THMB ABD	THMB ADD	THMB EXT	THMB FLX	REST		
	IND EXT	IND FLX	MID EXT	MID FLX	RING EXT	RING FLX	LIT EXT	LIT FLX	THMB ABD	THMB ADD	THMB EXT	THMB FLX	REST		

Figure 8: Confusion Matrix showing breakdown of individual finger movement classification for best performing 2 Layer LSTM

Further analysis of the confusion matrix, shown in figure 8, shows where the confusion is occurring. A similar pattern across the results show that the main area of confusion is with the movements of the thumb. Thumb extension and

flexion movements demonstrate to be the most difficult movements to classify with 61.5% and 71.4%, respectively. These signals were confused with the thumb adduction and abduction movements by the LSTM network. A potential cause of this is due to the fact that not all the muscles that are responsible for thumb movements are found in the forearm where the signals were being detected from. There are also intrinsic muscles found at the base of the thumb that control various aspects of the fine motor controls that the thumb can perform. These confusions of the thumb movements are consistent throughout the networks, figure 9 shows the varying levels of classification accuracy achieved when classifying these movements.




Figure 9: Accuracy of each thumb movement across each 2 layer LSTM

The closest comparison that can be made is with [23] where an accuracy of 71.66% was reported when classifying the same 12 movement signals. This research has improved on that finding with an average accuracy of 86.76% and some individual networks achieving 90% accuracy.

VI. CONCLUSION

This paper has demonstrated that Myo gesture control armband in combination with an LSTM network can be effectively implemented as a method of sEMG signal classification. The results of the experiments carried out show that whilst finger movements can be successfully classified they are much more difficult to classify if compared to classification of gestures that involve larger movements of the hand that involve multiple fingers or movements of the arm and wrist.

Developing a network that can successfully classify sEMG signals will allow further advancements in robotic object manipulation. Using human demonstrators with wearable sEMG devices and in combination with LSTM classifiers can be used not only for dexterous control of anthropomorphic robotic hands but also can be used as a method of informing a robotic grasping system.

To enhance the generality of these networks the addition of data from a range of subjects will be carried out. Further analysis of LSTM network parameters and how they affect the classification accuracy of the trained networks is needed in order to optimize the performance i.e. mini batch size, max epochs, number of hidden units and addition of frequency domain features. With the thumb showing to be most

difficult to classify an investigation into networks that classify finger movement's vs thumb movements as well as introduction of additional sEMG collection technology that can focus on the muscles responsible for thumb movements that are found within the hand.

REFERENCES

- [1] R. Hodson, "A gripping problem," *Nature*, vol. 557, no. 7704, pp. S23–S25, 2018.
- [2] B. Hudgins, P. Parker, and R. N. Scott, "A New Strategy for Multifunction Myoelectric Control," *IEEE Trans. Biomed. Eng.*, vol. 40, no. 1, pp. 82–94, 1993.
- [3] M. Zardoshti-Kermani, B. C. Wheeler, K. Badie, and R. M. Hashemi, "EMG Feature Evaluation for Movement Control of Upper Extremity Prostheses," *IEEE Trans. Rehabil. Eng.*, vol. 3, no. 4, pp. 324–333, 1995.
- [4] F. Tenore, A. Ramos, A. Fahmy, S. Acharya, R. Etienne-Cummings, and N. V. Thakor, "Towards the Control of Individual Fingers of a Prosthetic Hand Using Surface EMG Signals," in *2007 29th Annual International Conference of the IEEE Engineering in Medicine and Biology Society*, 2007, pp. 6145–6148.
- [5] A. Phinyomark, S. Hirunviriyaya, C. Limsakul, and P. Phukpattaranont, "Evaluation of EMG feature extraction for hand movement recognition based on Euclidean distance and standard deviation," *Electr. Eng. Comput. Telecommun. Inf. Technol. (ECTI-CON)*, 2010, no. May, pp. 856–860, 2010.
- [6] L. J. Hargrove, K. Englehart, and B. Hudgins, "A Comparison of Surface and Intramuscular Myoelectric Signal Classification," *IEEE Trans. Biomed. Eng.*, vol. 54, no. 5, pp. 847–853, May 2007.
- [7] F. V. G. Tenore, A. Ramos, A. Fahmy, S. Acharya, R. Etienne-Cummings, and N. V. Thakor, "Decoding of Individuated Finger Movements Using Surface Electromyography," *IEEE Trans. Biomed. Eng.*, vol. 56, no. 5, pp. 1427–1434, May 2009.
- [8] R. M. Stephenson, R. Chai, and D. Eager, "Isometric Finger Pose Recognition with Sparse Channel SpatioTemporal EMG Imaging," *Proc. Annu. Int. Conf. IEEE Eng. Med. Biol. Soc. EMBS*, vol. 2018-July, pp. 5232–5235, 2018.
- [9] M. Atzori *et al.*, "Electromyography data for non-invasive naturally-controlled robotic hand prostheses," *Sci. Data*, vol. 1, pp. 1–13, 2014.
- [10] A. Phinyomark, P. Phukpattaranont, and C. Limsakul, "Feature reduction and selection for EMG signal classification," *Expert Syst. Appl.*, vol. 39, no. 8, pp. 7420–7431, Jun. 2012.
- [11] M. R. Ahsan, M. I. Ibrahimy, and O. O. Khalifa, "Electromyography (EMG) signal based hand gesture recognition using artificial neural network (ANN)," *2011 4th Int. Conf. Mechatronics Integr. Eng. Ind. Soc. Dev. ICOM'11 - Conf. Proc.*, no. May, pp. 1–6, 2011.
- [12] S. Benatti *et al.*, "A Versatile Embedded Platform for EMG Acquisition and Gesture Recognition," *IEEE Trans. Biomed. Circuits Syst.*, vol. 9, no. 5, pp. 620–630, 2015.
- [13] A. Phinyomark and E. Scheme, "A feature extraction issue for myoelectric control based on wearable EMG sensors," in *2018 IEEE Sensors Applications Symposium (SAS)*, 2018, pp. 1–6.

- [14] I. Mendez *et al.*, “Evaluation of the Myo armband for the classification of hand motions,” *IEEE Int. Conf. Rehabil. Robot.*, pp. 1211–1214, 2017.
- [15] M. Atzori *et al.*, “A Benchmark Database for Myoelectric Movement Classification,” *Trans. Neural Syst. Rehabil. Eng.*, 2013.
- [16] K. Nymoen, M. Romarheim, H. Alexander, and R. Jensenius, “MuMYO — Evaluating and Exploring the MYO Armband for Musical Interaction,” *New Interfaces Music. Expr.*, p. Paper 179, 2015.
- [17] S. Shin, Y. Baek, J. Lee, Y. Eun, and S. H. Son, “Korean sign language recognition using EMG and IMU sensors based on group-dependent NN models,” *2017 IEEE Symp. Ser. Comput. Intell. SSCI 2017 - Proc.*, vol. 2018-Janua, pp. 1–7, 2018.
- [18] A. Phinyomark, C. Limsakul, and P. Phukpattaranont, “A Novel Feature Extraction for Robust EMG Pattern Recognition,” *J. Med. Eng. Technol.*, vol. 1, no. 1, pp. 71–80, Dec. 2009.
- [19] K. Tatarian, M. S. Couceiro, E. P. Ribeiro, and D. R. Faria, “Stepping-stones to Transhumanism: An EMG-controlled Low-cost Prosthetic Hand for Academia,” no. November, pp. 807–812, 2019.
- [20] M. E. Benalcazar, A. G. Jaramillo, Jonathan, A. Zea, A. Paez, and V. H. Andaluz, “Hand gesture recognition using machine learning and the Myo armband,” in *2017 25th European Signal Processing Conference (EUSIPCO)*, 2017, vol. 2017-Janua, pp. 1040–1044.
- [21] M. E. Benalcazar *et al.*, “Real-time hand gesture recognition using the Myo armband and muscle activity detection,” *2017 IEEE 2nd Ecuador Tech. Chapters Meet. ETCM 2017*, vol. 2017-Janua, pp. 1–6, 2018.
- [22] S. R. Kurniawan and D. Pamungkas, “MYO Armband sensors and Neural Network Algorithm for Controlling Hand Robot,” *Proc. 2018 Int. Conf. Appl. Eng. ICAE 2018*, pp. 1–6, 2018.
- [23] Y. Wu, B. Zheng, and Y. Zhao, “Dynamic Gesture Recognition Based on LSTM-CNN,” *Proc. 2018 Chinese Autom. Congr. CAC 2018*, pp. 2446–2450, 2019.
- [24] K. Akhmadeev, E. Rampone, T. Yu, Y. Aoustin, and E. Le Carpentier, “A testing system for a real-time gesture classification using surface EMG,” *IFAC-PapersOnLine*, vol. 50, no. 1, pp. 11498–11503, 2017.
- [25] P. Paudyal, J. Lee, A. Banerjee, and S. K. S. Gupta, “A comparison of techniques for sign language alphabet recognition using armband wearables,” *ACM Trans. Interact. Intell. Syst.*, vol. 9, no. 2–3, 2019.
- [26] R. N. Khushaba, A. Al-Ani, and A. Al-Jumaily, “Orthogonal Fuzzy Neighborhood Discriminant Analysis for Multifunction Myoelectric Hand Control,” *IEEE Trans. Biomed. Eng.*, vol. 57, no. 6, pp. 1410–1419, Jun. 2010.
- [27] M. Khezri and M. Jahed, “A neuro-fuzzy inference system for sEMG-based identification of hand motion commands,” *IEEE Trans. Ind. Electron.*, vol. 58, no. 5, pp. 1952–1960, 2011.
- [28] A. D. C. Chan and G. C. Green, “Myoelectric control development toolbox,” *Proc. 30th Conf. Can. Med. Biol. Eng. Soc.*, vol. 1, no. September, pp. M0100-1, 2007.
- [29] A. H. Al-Timemy, G. Bugmann, J. Escudero, and N. Outram, “Classification of finger movements for the dexterous hand prosthesis control with surface electromyography,” *IEEE J. Biomed. Heal. Informatics*, vol. 17, no. 3, pp. 608–618, 2013.
- [30] F. Quivira, T. Koike-Akino, Y. Wang, and D. Erdogmus, “Translating sEMG signals to continuous hand poses using recurrent neural networks,” *2018 IEEE EMBS Int. Conf. Biomed. Heal. Informatics, BHI 2018*, vol. 2018-Janua, no. March, pp. 166–169, 2018.
- [31] A. Gijberts, M. Atzori, C. Castellini, H. Müller, and B. Caputo, “Movement error rate for evaluation of machine learning methods for sEMG-based hand movement classification,” *IEEE Trans. Neural Syst. Rehabil. Eng.*, vol. 22, no. 4, pp. 735–744, 2014.
- [32] S. Pizzolato, L. Tagliapietra, M. Cognolato, M. Reggiani, H. Müller, and M. Atzori, “Comparison of six electromyography acquisition setups on hand movement classification tasks,” *PLoS One*, vol. 12, no. 10, pp. 1–17, 2017.
- [33] S. Hochreiter and J. Urgan Schmidhuber, “Long Shortterm Memory,” *Neural Comput.*, vol. 9, no. 8, p. 17351780, 1997.
- [34] T. Cooijmans, N. Ballas, C. Laurent, Ç. Gülçehre, and A. Courville, “Recurrent batch normalization,” in *5th International Conference on Learning Representations, ICLR 2017 - Conference Track Proceedings*, 2019, no. Section 3, pp. 1–13.
- [35] M. Tomaszewski, “Myo SDK MATLAB MEX Wrapper,” 2019.
- [36] M. Dwarampudi and N. V. S. Reddy, “Effects of padding on LSTMs and CNNs,” Mar. 2019.